



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB101373

Analyst By : AMANDEEP

Supervisor Review By : apatel

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/12/2019	09:45
2	CAL2	1	Water	NA	NA	20.2	7.02	03/12/2019	09:47
3	CAL3	1	Water	NA	NA	20.2	10.02	03/12/2019	09:48
4	ICV	1	Water	NA	NA	20.1	7.01	03/12/2019	09:49
5	CCV1	1	Water	NA	NA	20.3	2.02	03/12/2019	09:50
6	K1844-03	1	Water	NA	NA	14.6	7.60	03/12/2019	10:39
7	K1844-03DUP	1	Water	NA	NA	14.6	7.60	03/12/2019	10:40
8	CCV2	1	Water	NA	NA	20.2	12.01	03/12/2019	10:42

WORKLIST(Hardcopy Internal Chain)

B101373

WorkList Name : corrosivity031219-2

WorkList ID : 123301

Date : 03-12-2019 11:11:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/11/2019	Water	K1844-03	Corrosivity	Cool 4 deg C	PSEG01	K31	DRUM-1-4	03/11/2019	9040C

Date/Time 3-12-19 9:20
 Received by: AS
 Relinquished by: AS

Date/Time 3-12-19 11:00
 Received by: CS
 Relinquished by: no