



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101384

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/14/2019	16:45
2	CAL2	1	Water	NA	NA	20.2	7.01	03/14/2019	16:46
3	CAL3	1	Water	NA	NA	20.2	10.02	03/14/2019	16:47
4	ICV	1	Water	NA	NA	20.2	7.00	03/15/2019	16:48
5	CCV1	1	Water	NA	NA	20.2	2.01	03/14/2019	16:49
6	K1923-01	1	Water	NA	NA	19.1	4.66	03/14/2019	16:50
7	K1923-01DUP	1	Water	NA	NA	19.2	4.67	03/14/2019	16:51
8	K1923-02	1	Water	NA	NA	19.3	4.97	03/14/2019	16:52
9	K1925-01	1	Water	NA	NA	17.7	4.83	03/14/2019	17:00
10	CCV2	1	Water	NA	NA	20.2	12.01	03/14/2019	17:01

LB 101384

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1823

WorkList ID : 123437

Date : 03-14-2019 16:35:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/13/2019	Water	K1923-01 - A	pH	Cool 4 deg C	STER05	P33	BOTTLE-1	03/13/2019	SM 4500-H B
03/13/2019	Water	K1923-02 - A	pH	Cool 4 deg C	STER05	P33	BOTTLE-2	03/13/2019	SM 4500-H B
03/08/2019	Water	K1925-01 - A	pH	Cool 4 deg C	STER14	P31	JS-SYSTEM	03/08/2019	SM 4500-H B

Date/Time 03-14-19 16:40
 Received by: Jn
 Relinquished by: Jn

Date/Time 03-14-19 17:30
 Received by: Jn
 Relinquished by: Jn