


LB101389

7470A

INSTRUMENT ID: CUI

 μ Abs.:

43401



Linear

A= 0.0000e+000

B= 2.2995e-004

C= 2.1506e-002

Rho= 0.9999302

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	18	0.000	18				
0.2	0.200	0.215	0.015	840	0.0 %	840				
2.5	2.500	2.485	-0.015	10714	0.0 %	10714				
5.0	5.000	4.916	-0.084	21285	0.0 %	21285				
7.5	7.500	7.557	0.057	32770	0.0 %	32770				
10.0	10.000	10.002	0.002	43401	0.0 %	43401				

LB101389 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Mar 2019 09:45:38

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50.0	18	-	0.00007470A	PPB	15 Mar 2019 09:55:26	S	Std
0.2 - 1	50.2	840	-	0.20007470A	PPB	15 Mar 2019 09:57:35	S	Std
2.5 - 1	52.5	10714	-	2.50007470A	PPB	15 Mar 2019 09:59:43	S	Std
5.0 - 1	55.0	21285	-	5.00007470A	PPB	15 Mar 2019 10:01:51	S	Std
7.5 - 1	57.5	32770	-	7.50007470A	PPB	15 Mar 2019 10:03:58	S	Std
10.0 - 1	510.0	43401	-	10.00007470A	PPB	15 Mar 2019 10:06:04	S	Std
ICV55 - 1	ICV55	16363	3.7842	-7470A	PPB	15 Mar 2019 10:12:05	U	SMPL
ICB55 - 1	ICB55	-8	0.0197	-7470A	PPB	15 Mar 2019 10:14:11	U	SMPL
CCV67 - 1	CCV67	21176	4.8909	-7470A	PPB	15 Mar 2019 10:16:17	U	SMPL
CCB67 - 1	CCB67	58	0.0348	-7470A	PPB	15 Mar 2019 10:18:23	U	SMPL
CRA - 1	CRA	808	0.2073	-7470A	PPB	15 Mar 2019 10:20:29	U	SMPL
HighStd - 1	HighStd	43586	10.0441	-7470A	PPB	15 Mar 2019 10:22:35	U	SMPL
ChkStd - 1	ChkStd	31331	7.2261	-7470A	PPB	15 Mar 2019 10:24:41	U	SMPL
PB117949BL - 1	PBW	17	0.0254	-7470A	PPB	15 Mar 2019 10:29:17	U	SMPL
PB117949BS - 1	LCSW	17578	4.0636	-7470A	PPB	15 Mar 2019 10:31:23	U	SMPL
K1882-04 - 1	BPS1-TT-MW305S-20190312-F	31	0.0286	-7470A	PPB	15 Mar 2019 10:33:31	U	SMPL
K1882-06 - 1	BP-DUP01-20190312-F	80	0.0399	-7470A	PPB	15 Mar 2019 10:35:38	U	SMPL
K1900-02 - 1	BPS1-TT-MW304I2-20190312	91	0.0424	-7470A	PPB	15 Mar 2019 10:37:46	U	SMPL
K1900-03 - 1	BPS1-TT-MW304I1-20190312	89	0.0420	-7470A	PPB	15 Mar 2019 10:39:54	U	SMPL
K1900-04 - 1	BP-HN-MW27I-20190312	220	0.0721	-7470A	PPB	15 Mar 2019 10:42:00	U	SMPL
CCV68 - 1	CCV68	21913	5.0604	-7470A	PPB	15 Mar 2019 10:44:06	U	SMPL
CCB68 - 1	CCB68	-13	0.0185	-7470A	PPB	15 Mar 2019 10:46:12	U	SMPL
K1900-05 - 1	BP-TT-AOC22-MW06-20190312	451	0.1252	-7470A	PPB	15 Mar 2019 10:48:18	U	SMPL
K1900-06 - 1	BP-TT-AOC22-MW06-20190312-F	96	0.0436	-7470A	PPB	15 Mar 2019 10:50:24	U	SMPL
K1900-07 - 1	BPS1-TT-MW309D-20190313	89	0.0420	-7470A	PPB	15 Mar 2019 10:52:30	U	SMPL
K1900-07DUP - 1	BPS1-TT-MW309D-20190313DUP	91	0.0424	-7470A	PPB	15 Mar 2019 10:54:37	U	SMPL
K1900-08 - 1	K1900-07MS	17376	4.0171	-7470A	PPB	15 Mar 2019 10:56:43	U	SMPL
K1900-09 - 1	K1900-07MSD	17462	4.0369	-7470A	PPB	15 Mar 2019 10:58:50	U	SMPL
K1900-10 - 1	BPS1-TT-MW309D-20190313-F	311	0.0930	-7470A	PPB	15 Mar 2019 11:00:58	U	SMPL
K1900-11 - 1	BP-DUP02-20190313	61	0.0355	-7470A	PPB	15 Mar 2019 11:03:06	U	SMPL
K1900-12 - 1	BP-DUP02-20190313-F	80	0.0399	-7470A	PPB	15 Mar 2019 11:05:13	U	SMPL
K1900-13 - 1	BPS1-TT-MW305I-20190313	92	0.0427	-7470A	PPB	15 Mar 2019 11:07:20	U	SMPL
CCV69 - 1	CCV69	21828	5.0408	-7470A	PPB	15 Mar 2019 11:09:26	U	SMPL
CCB69 - 1	CCB69	-12	0.0187	-7470A	PPB	15 Mar 2019 11:11:32	U	SMPL
K1900-14 - 1	BPS1-TT-MW309S-20190313	87	0.0415	-7470A	PPB	15 Mar 2019 11:13:39	U	SMPL
K1900-15 - 1	BPS1-TT-MW301S-20190313	78	0.0394	-7470A	PPB	15 Mar 2019 11:15:45	U	SMPL
PB117950BL - 1	PBW	13	0.0245	-7470A	PPB	15 Mar 2019 11:17:52	U	SMPL
PB117950BS - 1	LCSW	17614	4.0718	-7470A	PPB	15 Mar 2019 11:19:59	U	SMPL
K1902-01 - 1	51064	150	0.0560	-7470A	PPB	15 Mar 2019 11:22:06	U	SMPL
K1902-01DUP - 1	51064DUP	197	0.0668	-7470A	PPB	15 Mar 2019 11:24:13	U	SMPL
K1902-01MS - 1	51064MS	17459	4.0362	-7470A	PPB	15 Mar 2019 11:26:20	U	SMPL
K1902-01MSD - 1	51064MSD	17650	4.0801	-7470A	PPB	15 Mar 2019 11:28:28	U	SMPL
K1902-02 - 1	30276	350	0.1020	-7470A	PPB	15 Mar 2019 11:30:36	U	SMPL
K1902-03 - 1	30298	187	0.0645	-7470A	PPB	15 Mar 2019 11:32:43	U	SMPL
CCV70 - 1	CCV70	21283	4.9155	-7470A	PPB	15 Mar 2019 11:34:51	U	SMPL
CCB70 - 1	CCB70	-11	0.0190	-7470A	PPB	15 Mar 2019 11:37:02	U	SMPL
K1902-04 - 1	30299	182	0.0634	-7470A	PPB	15 Mar 2019 11:39:10	U	SMPL
K1902-05 - 1	30300	192	0.0657	-7470A	PPB	15 Mar 2019 11:41:17	U	SMPL
K1902-06 - 1	30301	190	0.0652	-7470A	PPB	15 Mar 2019 11:43:23	U	SMPL
K1902-07 - 1	30302	203	0.0682	-7470A	PPB	15 Mar 2019 11:45:31	U	SMPL
K1902-08 - 1	30303	204	0.0684	-7470A	PPB	15 Mar 2019 11:47:37	U	SMPL
K1902-09 - 1	30282	169	0.0604	-7470A	PPB	15 Mar 2019 11:49:45	U	SMPL
K1902-10 - 1	51065	227	0.0737	-7470A	PPB	15 Mar 2019 11:51:54	U	SMPL
K1902-12 - 1	30240-41-42-43-44	269	0.0834	-7470A	PPB	15 Mar 2019 11:54:03	U	SMPL
K1902-14 - 1	30296-97-30304-05-06	233	0.0751	-7470A	PPB	15 Mar 2019 11:56:11	U	SMPL
K1902-16 - 1	30307-08-09-10-11	357	0.1036	-7470A	PPB	15 Mar 2019 11:58:18	U	SMPL
CCV71 - 1	CCV71	21373	4.9362	-7470A	PPB	15 Mar 2019 12:00:25	U	SMPL
CCB71 - 1	CCB71	-29	0.0148	-7470A	PPB	15 Mar 2019 12:02:31	U	SMPL
K1902-18 - 1	30312-13-14-15-17	557	0.1496	-7470A	PPB	15 Mar 2019 12:04:38	U	SMPL
PB117951BL - 1	PBW	11	0.0240	-7470A	PPB	15 Mar 2019 12:06:44	U	SMPL
PB117951BS - 1	LCSW	17499	4.0454	-7470A	PPB	15 Mar 2019 12:08:51	U	SMPL
K1866-02 - 1	MR-MH-04-COMP	239	0.0765	-7470A	PPB	15 Mar 2019 12:10:57	U	SMPL
K1866-02DUP - 1	MR-MH-04-COMPDUP	257	0.0806	-7470A	PPB	15 Mar 2019 12:13:04	U	SMPL
K1866-02MS - 1	MR-MH-04-COMPMS	17448	4.0337	-7470A	PPB	15 Mar 2019 12:15:11	U	SMPL
K1866-02MSD - 1	MR-MH-04-COMPMSD	17126	3.9596	-7470A	PPB	15 Mar 2019 12:17:18	U	SMPL
K1866-04 - 1	MR-MH-03-COMP	189	0.0650	-7470A	PPB	15 Mar 2019 12:19:26	U	SMPL
K1866-06 - 1	MR-MH-02-COMP	265	0.0824	-7470A	PPB	15 Mar 2019 12:21:34	U	SMPL
K1866-08 - 1	MR-MH-01-COMP	282	0.0864	-7470A	PPB	15 Mar 2019 12:23:41	U	SMPL
CCV72 - 1	CCV72	22041	5.0898	-7470A	PPB	15 Mar 2019 12:25:48	U	SMPL
CCB72 - 1	CCB72	-19	0.0171	-7470A	PPB	15 Mar 2019 12:27:56	U	SMPL
K1866-10 - 1	MR-MH-15-COMP	284	0.0868	-7470A	PPB	15 Mar 2019 12:30:03	U	SMPL
K1866-12 - 1	MR-MH-14-COMP	245	0.0778	-7470A	PPB	15 Mar 2019 12:32:09	U	SMPL
K1866-14 - 1	MR-MH-13-COMP	205	0.0686	-7470A	PPB	15 Mar 2019 12:34:16	U	SMPL
K1866-16 - 1	MR-MH-12-COMP	229	0.0742	-7470A	PPB	15 Mar 2019 12:36:23	U	SMPL
K1892-01 - 1	GARAGE-FLOOR	215	0.0709	-7470A	PPB	15 Mar 2019 12:38:30	U	SMPL
K1895-02 - 1	OILY-SOIL	202	0.0680	-7470A	PPB	15 Mar 2019 12:40:37	U	SMPL
K1897-04 - 1	TP-1	191	0.0654	-7470A	PPB	15 Mar 2019 12:42:45	U	SMPL
K1897-09 - 1	TP-2	225	0.0732	-7470A	PPB	15 Mar 2019 12:44:53	U	SMPL
K1902-11 - 1	30245	164	0.0592	-7470A	PPB	15 Mar 2019 12:47:01	U	SMPL
PB117950TB - 1	PB117900TB	47	0.0323	-7470A	PPB	15 Mar 2019 12:49:08	U	SMPL
CCV73 - 1	CCV73	21788	5.0317	-7470A	PPB	15 Mar 2019 12:51:15	U	SMPL
CCB73 - 1	CCB73	-14	0.0183	-7470A	PPB	15 Mar 2019 12:53:22	U	SMPL
PB117951TB - 1	PB117897TB	30	0.0284	-7470A	PPB	15 Mar 2019 12:55:29	U	SMPL
K1900-07LX5 - 1		25	0.0273	-7470A	PPB	15 Mar 2019 12:57:37	U	SMPL
K1900-07A - 1		17432	4.0300	-7470A	PPB	15 Mar 2019 12:59:44	U	SMPL
K1902-01LX5 - 1		20	0.0261	-7470A	PPB	15 Mar 2019 13:01:51	U	SMPL
K1902-01A - 1		17266	3.9918	-7470A	PPB	15 Mar 2019 13:03:58	U	SMPL
K1866-02LX5 - 1		23	0.0268	-7470A	PPB	15 Mar 2019 13:06:05	U	SMPL
K1866-02A - 1		17292	3.9978	-7470A	PPB	15 Mar 2019 13:08:12	U	SMPL
CCV74 - 1	CCV74	21434	4.9502	-7470A	PPB	15 Mar 2019 13:10:20	U	SMPL
CCB74 - 1	CCB74	-25	0.0158	-7470A	PPB	15 Mar 2019 13:12:28	U	SMPL