


7471B

LB101401
INSTRUMENT ID: CV1

Linear

 μ Abs.:

60103



A= 0.0000e+000

B= 1.6779e-004

C= -8.7757e-002

Rho= 0.9996998

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.087	-0.087	3	0.000	3				
0.2	0.200	0.182	-0.018	1605	0.0 %	1605				
2.5	2.500	2.603	0.103	16037	0.0 %	16037				
5.0	5.000	5.126	0.126	31073	0.0 %	31073				
7.5	7.500	7.379	-0.121	44502	0.0 %	44502				
10.0	10.000	9.997	-0.003	60103	0.0 %	60103				

LB101401 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Mar 2019 13:54:18

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO-0	3	-	0.0000	7471B	PPB	15 Mar 2019 13:58:32	S	Std
0.2 - 1	SO-2	1605	-	0.2000	7471B	PPB	15 Mar 2019 14:00:49	S	Std
2.5 - 1	SO-5	16037	-	2.5000	7471B	PPB	15 Mar 2019 14:03:06	S	Std
5.0 - 1	SO-10	31073	-	5.0000	7471B	PPB	15 Mar 2019 14:05:23	S	Std
7.5 - 1	SO-15	44502	-	7.5000	7471B	PPB	15 Mar 2019 14:07:40	S	Std
10.0 - 1	SO-20	60103	-	10.0000	7471B	PPB	15 Mar 2019 14:19:49	S	Std
ICV56 - 1	ICV56	24646	4.0477	-	7471B	PPB	15 Mar 2019 14:24:50	U	SMPL
ICB56 - 1	ICB56	273	-0.0419	-	7471B	PPB	15 Mar 2019 14:27:05	U	SMPL
CCV75 - 1	CCV75	31201	5.1476	-	7471B	PPB	15 Mar 2019 14:29:20	U	SMPL
CCB75 - 1	CCB75	-58	-0.0975	-	7471B	PPB	15 Mar 2019 14:31:36	U	SMPL
CRA - 1	CRA	1588	0.1787	-	7471B	PPB	15 Mar 2019 14:33:51	U	SMPL
HighStd - 1	HighStd	60380	10.0436	-	7471B	PPB	15 Mar 2019 14:36:06	U	SMPL
ChkStd - 1	ChkStd	41839	6.9325	-	7471B	PPB	15 Mar 2019 14:38:21	U	SMPL
PB117944BL - 1	PBS	76	-0.0750	-	7471B	PPB	15 Mar 2019 14:40:39	U	SMPL
PB117944BS - 1	LCSS	22439	3.6774	-	7471B	PPB	15 Mar 2019 14:42:56	U	SMPL
K1895-01 - 1	OILY-SOIL	2291	0.2967	-	7471B	PPB	15 Mar 2019 14:45:13	U	SMPL
K1895-01DUP - 1	OILY-SOILDUP	2302	0.2985	-	7471B	PPB	15 Mar 2019 14:47:30	U	SMPL
K1895-01MS - 1	OILY-SOILMS	23095	3.7874	-	7471B	PPB	15 Mar 2019 14:49:47	U	SMPL
K1895-01MSD - 1	OILY-SOILMSD	23264	3.8158	-	7471B	PPB	15 Mar 2019 14:52:04	U	SMPL
K1897-01 - 1	TP-1	3508	0.5009	-	7471B	PPB	15 Mar 2019 14:54:20	U	SMPL
CCV76 - 1	CCV76	31325	5.1684	-	7471B	PPB	15 Mar 2019 14:56:36	U	SMPL
CCB76 - 1	CCB76	68	-0.0763	-	7471B	PPB	15 Mar 2019 14:58:51	U	SMPL
K1897-06 - 1	TP-2	3772	0.5452	-	7471B	PPB	15 Mar 2019 15:01:08	U	SMPL
K1901-01 - 1	SB-1(0.5-1.5)	4379	0.6470	-	7471B	PPB	15 Mar 2019 15:03:23	U	SMPL
K1901-02 - 1	SB-1(12-13)	11345	1.8159	-	7471B	PPB	15 Mar 2019 15:05:39	U	SMPL
K1901-03 - 1	SB-3(0.5-1.5)	59198	9.8453	-	7471B	PPB	15 Mar 2019 15:07:54	U	SMPL
K1901-04 - 1	SB-3(7-8)	57247	9.5179	-	7471B	PPB	15 Mar 2019 15:10:10	U	SMPL
K1901-05 - 1	SB-4(0.5-1.5)	31160	5.1407	-	7471B	PPB	15 Mar 2019 15:12:28	U	SMPL
K1901-06 - 1	SB-4(7-8)	8578	1.3516	-	7471B	PPB	15 Mar 2019 15:14:46	U	SMPL
K1901-07 - 1	SB-6(0.5-1.5)	59796	9.9456	-	7471B	PPB	15 Mar 2019 15:17:04	U	SMPL
K1901-08 - 1	SB-6(6.5-7.5)	294620	49.3474	-	7471B	PPB	15 Mar 2019 15:19:21	U	SMPL
K1901-09 - 1	SB-16(0.5-1.5)	31354	5.1732	-	7471B	PPB	15 Mar 2019 15:21:38	U	SMPL
CCV77 - 1	CCV77	31900	5.2648	-	7471B	PPB	15 Mar 2019 15:23:58	U	SMPL
CCB77 - 1	CCB77	51	-0.0792	-	7471B	PPB	15 Mar 2019 15:26:14	U	SMPL
K1901-10 - 1	SB-16(7-8)	90189	15.0453	-	7471B	PPB	15 Mar 2019 15:28:30	U	SMPL
K1905-01 - 1	RT5124	1796	0.2136	-	7471B	PPB	15 Mar 2019 15:30:46	U	SMPL
K1905-03 - 1	RBR-200052	1790	0.2126	-	7471B	PPB	15 Mar 2019 15:33:03	U	SMPL
K1905-05 - 1	RBR-200050	1890	0.2294	-	7471B	PPB	15 Mar 2019 15:35:19	U	SMPL
K1905-07 - 1	RT-3483	4348	0.6418	-	7471B	PPB	15 Mar 2019 15:37:35	U	SMPL
K1908-01 - 1	MH-02	7725	1.2084	-	7471B	PPB	15 Mar 2019 15:39:52	U	SMPL
K1908-03 - 1	101-102	4730	0.7059	-	7471B	PPB	15 Mar 2019 15:42:08	U	SMPL
K1908-05 - 1	93-94	1607	0.1819	-	7471B	PPB	15 Mar 2019 15:44:26	U	SMPL
PB117946BL - 1	PBS	177	-0.0581	-	7471B	PPB	15 Mar 2019 15:46:43	U	SMPL
PB117946BS - 1	LCSS	22679	3.7176	-	7471B	PPB	15 Mar 2019 15:48:59	U	SMPL
CCV78 - 1	CCV78	31686	5.2289	-	7471B	PPB	15 Mar 2019 15:51:15	U	SMPL
CCB78 - 1	CCB78	230	-0.0492	-	7471B	PPB	15 Mar 2019 15:53:31	U	SMPL
K1685-01 - 1	AU-05-031419-A	3811	0.5517	-	7471B	PPB	15 Mar 2019 15:55:48	U	SMPL
K1685-03 - 1	AU-05-031419-B	3865	0.5608	-	7471B	PPB	15 Mar 2019 15:58:05	U	SMPL
K1685-05 - 1	AU-05-031419-C	2616	0.3512	-	7471B	PPB	15 Mar 2019 16:00:21	U	SMPL
K1685-07 - 1	AU-05-031419-D	3149	0.4406	-	7471B	PPB	15 Mar 2019 16:02:37	U	SMPL
K1698-01 - 1	TR-04-031419	1708	0.1988	-	7471B	PPB	15 Mar 2019 16:04:53	U	SMPL
K1906-01 - 1	OILY-DEBRIS	165	-0.0601	-	7471B	PPB	15 Mar 2019 16:07:09	U	SMPL
K1911-01 - 1	RBR200036	1042	0.0871	-	7471B	PPB	15 Mar 2019 16:09:26	U	SMPL
K1911-04 - 1	RBR200051	2334	0.3039	-	7471B	PPB	15 Mar 2019 16:11:43	U	SMPL
K1914-01 - 1	CONCRETE-PILE	252	-0.0455	-	7471B	PPB	15 Mar 2019 16:14:00	U	SMPL
K1921-01 - 1	282-346	13103	2.1108	-	7471B	PPB	15 Mar 2019 16:16:16	U	SMPL
CCV79 - 1	CCV79	32104	5.2991	-	7471B	PPB	15 Mar 2019 16:18:32	U	SMPL
CCB79 - 1	CCB79	224	-0.0502	-	7471B	PPB	15 Mar 2019 16:20:48	U	SMPL
K1922-01 - 1	ORG-031419	2628	0.3532	-	7471B	PPB	15 Mar 2019 16:23:05	U	SMPL
K1924-01 - 1	NB-SOIL-PILE-A	1646	0.1884	-	7471B	PPB	15 Mar 2019 16:25:21	U	SMPL
K1924-03 - 1	NB-SOIL-PILE-B	1266	0.1247	-	7471B	PPB	15 Mar 2019 16:27:36	U	SMPL
K1927-01 - 1	DRUM-COMP	260	-0.0441	-	7471B	PPB	15 Mar 2019 16:29:52	U	SMPL
K1930-01 - 1	TP-3	3144	0.4398	-	7471B	PPB	15 Mar 2019 16:32:09	U	SMPL
K1930-05 - 1	TP-4	3161	0.4426	-	7471B	PPB	15 Mar 2019 16:34:25	U	SMPL
K1930-10 - 1	TP-5	1005	0.0809	-	7471B	PPB	15 Mar 2019 16:36:42	U	SMPL
K1930-14 - 1	TP-6	789	0.0446	-	7471B	PPB	15 Mar 2019 16:39:00	U	SMPL
K1930-14DUP - 1	TP-6DUP	809	0.0480	-	7471B	PPB	15 Mar 2019 16:41:20	U	SMPL
K1930-14MS - 1	TP-6MS	23171	3.8002	-	7471B	PPB	15 Mar 2019 16:43:36	U	SMPL
CCV80 - 1	CCV80	32022	5.2853	-	7471B	PPB	15 Mar 2019 16:45:52	U	SMPL
CCB80 - 1	CCB80	279	-0.0409	-	7471B	PPB	15 Mar 2019 16:48:09	U	SMPL
K1930-14MSD - 1	TP-6MSD	23058	3.7812	-	7471B	PPB	15 Mar 2019 16:50:25	U	SMPL
K1895-01LX5 - 1		439	-0.0141	-	7471B	PPB	15 Mar 2019 16:52:41	U	SMPL
K1895-01A - 1		23561	3.8656	-	7471B	PPB	15 Mar 2019 16:54:58	U	SMPL
K1930-14LX5 - 1		116	-0.0683	-	7471B	PPB	15 Mar 2019 16:57:14	U	SMPL
K1930-14A - 1		23026	3.7758	-	7471B	PPB	15 Mar 2019 16:59:30	U	SMPL
K1901-08DLX10 - 1		32293	5.3308	-	7471B	PPB	15 Mar 2019 17:01:47	U	SMPL
K1901-10DLX2 - 1		45369	7.5248	-	7471B	PPB	15 Mar 2019 17:04:03	U	SMPL
CCV81 - 1	CCV81	32216	5.3179	-	7471B	PPB	15 Mar 2019 17:06:21	U	SMPL
CCB81 - 1	CCB81	240	-0.0475	-	7471B	PPB	15 Mar 2019 17:08:39	U	SMPL