



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101408

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.2	03/15/2019	09:50
2	K1909-01		1	9	>100.0	>212.0	>212.0	03/15/2019	10:20
3	K1909-01DUP		1	9	>100.0	>212.0	>212.0	03/15/2019	10:45
4	K1909-02		1	7	>100.0	>212.0	>212.0	03/15/2019	11:10
5	K1909-03		1	8	>100.0	>212.0	>212.0	03/15/2019	11:40
6	K1909-04		1	8	>100.0	>212.0	>212.0	03/15/2019	12:05
7	K1909-05		1	9	>100.0	>212.0	>212.0	03/15/2019	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB101408

WorkList Name : FP31519

WorkList ID : 123483

Date : 03-15-2019 09:20:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/28/2019	Water	K1909-01 H	Flash Point	Cool 4 deg C	TULL02	P41	MH-2	03/14/2019	1010A
03/28/2019	Water	K1909-02 G	Flash Point	Cool 4 deg C	TULL02	P41	101-102	03/14/2019	1010A
03/28/2019	Water	K1909-03 G	Flash Point	Cool 4 deg C	TULL02	P41	93-94	03/14/2019	1010A
03/28/2019	Water	K1909-04 F	Flash Point	Cool 4 deg C	TULL02	P41	97-98	03/14/2019	1010A
03/28/2019	Water	K1909-05 G	Flash Point	Cool 4 deg C	TULL02	P41	MH-CO5(SETTLE-WATER)	03/14/2019	1010A

Date/Time 03/15/19 09:25
 Received by: KS
 Relinquished by: CP

Date/Time 03/15/19 12:45
 Received by: CP
 Relinquished by: KS