

LB101415


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42162



A= 0.0000e+000

B= 2.3819e-004

C= -6.9000e-002

Rho= 0.9999053

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.062	-0.062	31	0.000	31					
0.2	0.200	0.181	-0.019	1048	0.0 %	1048					-10
2.5	2.500	2.600	0.100	11204	0.0 %	11204					4
5.0	5.000	5.022	0.022	21373	0.0 %	21373					000
7.5	7.500	7.486	-0.014	31719	0.0 %	31719					
10.0	10.000	9.973	-0.027	42162	0.0 %	42162					

LB101415

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	31	-		0 ISM02.4	PPB	3/15/2019 17:24	Std
	0.2 S0.2	1048	-		0.2 ISM02.4	PPB	3/15/2019 17:26	Std
	2.5 S2.5	11204	-		2.5 ISM02.4	PPB	3/15/2019 17:29	Std
	5 S5.0	21373	-		5 ISM02.4	PPB	3/15/2019 17:31	Std
	7.5 S7.5	31719	-		7.5 ISM02.4	PPB	3/15/2019 17:33	Std
	10 S10.0	42162	-		10 ISM02.4	PPB	3/15/2019 17:35	Std
ICV	ICV	17222	4.0331 -		ISM02.4	PPB	3/15/2019 17:39	SMPL
ICB	ICB	148	-0.0337 -		ISM02.4	PPB	3/15/2019 17:41	SMPL
CCV082	CCV082	21502	5.0525 -		ISM02.4	PPB	3/15/2019 17:43	SMPL
CCB082	CCB082	164	-0.0299 -		ISM02.4	PPB	3/15/2019 17:45	SMPL
PB117982BL	PBW982	212	-0.0185 -		ISM02.4	PPB	3/15/2019 17:48	SMPL
K1774-02	MA41A3	16110	3.7682 -		ISM02.4	PPB	3/15/2019 18:04	SMPL
K1774-03	MA41D9	172	-0.028 -		ISM02.4	PPB	3/15/2019 18:06	SMPL
K1774-04	MA41E0	169	-0.0287 -		ISM02.4	PPB	3/15/2019 18:08	SMPL
K1774-05	MA41E8	162	-0.0304 -		ISM02.4	PPB	3/15/2019 18:11	SMPL
K1774-06	MA41F5	163	-0.0302 -		ISM02.4	PPB	3/15/2019 18:13	SMPL
K1774-07	MA41N1	169	-0.0287 -		ISM02.4	PPB	3/15/2019 18:15	SMPL
K1774-08	MA41E0D	226	-0.0152 -		ISM02.4	PPB	3/15/2019 18:17	SMPL
K1774-09	MA41E0S	4926	1.1043 -		ISM02.4	PPB	3/15/2019 18:20	SMPL
CCV083	CCV083	21511	5.0547 -		ISM02.4	PPB	3/15/2019 18:22	SMPL
CCB083	CCB083	-15	-0.0726 -		ISM02.4	PPB	3/15/2019 18:24	SMPL