



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101428

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/18/2019	15:00
2	CAL2	1	Water	NA	NA	20.2	7.01	03/18/2019	15:01
3	CAL3	1	Water	NA	NA	20.2	10.04	03/18/2019	15:02
4	ICV	1	Water	NA	NA	20.2	6.99	03/18/2019	15:04
5	CCV1	1	Water	NA	NA	20.2	2.01	03/18/2019	15:05
6	K1931-01	1	Solid	20.02	20	20.6	5.97	03/18/2019	15:06
7	K1974-01	1	Solid	20.04	20	19.9	7.24	03/18/2019	15:08
8	K1974-01DUP	1	Solid	20.05	20	19.8	7.25	03/18/2019	15:09
9	CCV2	1	Water	NA	NA	20.1	12.02	03/18/2019	15:10

28101428

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1974

WorkList ID : 123568

Date : 03-18-2019 14:49:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/13/2019	Solid	K1931-01-D	pH	Cool 4 deg C	AREC01	F11	LSP-1	03/13/2019	9045D
03/14/2019	Solid	K1974-01 ✓	pH	Cool 4 deg C	GENV01	F41	WC4	03/14/2019	9045D

Date/Time 03-18-19 15:00

Received by: J

Relinquished by: C

Date/Time

03-18-19 16:00

Received by: J

Relinquished by: C