



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101429

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/18/2019	15:00
2	CAL2	1	Water	NA	NA	20.2	7.01	03/18/2019	15:01
3	CAL3	1	Water	NA	NA	20.2	10.04	03/18/2019	15:02
4	ICV	1	Water	NA	NA	20.2	6.99	03/18/2019	15:04
5	CCV1	1	Water	NA	NA	20.1	2.02	03/18/2019	15:15
6	K1941-01	1	Water	NA	NA	18.5	7.53	03/18/2019	15:16
7	K1941-01DUP	1	Water	NA	NA	18.6	7.54	03/18/2019	15:17
8	K1941-04	1	Water	NA	NA	19.1	7.60	03/18/2019	15:19
9	CCV2	1	Water	NA	NA	20.2	12.01	03/18/2019	15:20

28101429

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1941

WorkList ID : 123570

Date : 03-18-2019 14:51:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/18/2019	Water-A	K1941-01	pH	Cool 4 deg C	PSEG01	K31	T-1	03/18/2019	9040C
03/18/2019	Water-A	K1941-04	pH	Cool 4 deg C	PSEG01	K31	T-2	03/18/2019	9040C

Date/Time 03-18-19 15:00

Received by: Jy

Relinquished by:

Date/Time 03-18-19 16:20

Received by: Jy

Relinquished by: Jy