



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/19/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 03/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 03/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101432

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1692-03	OK-02-031819	1	1.13	9.6	8.89	91.6
K1693-09	OR-03-031819-A	2	1.12	9.72	9.00	91.6
K1693-11	OR-03-031819-B	3	1.15	9.8	9.11	92.0
K1941-07	SP	4	1.11	9.79	8.71	87.6
K1941-08	SP	5	1.11	9.79	8.71	87.6
K1943-01	20190124-LOC-1	6	1.13	9.78	3.08	22.5
K1943-02	20190125-LOC-2	7	1.15	9.95	3.25	23.9
K1943-03	20190126-LOC-3	8	1.13	9.94	4.26	35.5
K1943-04	20190127-LOC-4	9	1.13	9.61	2.8	19.7
K1943-05	20190128-LOC-5	10	1.16	9.78	2.32	13.5
K1943-06	20190129-LOC-6	11	1.13	9.6	2.26	13.3
K1943-07	20190130-LOC-7	12	1.17	9.77	2.98	21.0
K1943-08	20190131-LOC-8	13	1.15	9.98	3.85	30.6
K1943-09	20190132-LOC-9	14	1.14	9.97	3.61	28.0
K1943-10	20190133-LOC-10	15	1.17	9.93	4.66	39.8
K1944-01	20190136-WIPE-1-BARRELS	16	1.00	2.00	2.00	100.0
K1944-02	20190137-WIPE-2-BOAT	17	1.00	2.00	2.00	100.0
K1944-03	20190138-WIPE-3-TOOLS	18	1.00	2.00	2.00	100.0
K1944-04	20190139-WIPE-4-CONES	19	1.00	2.00	2.00	100.0
K1944-05	20190140-WIPE-5-PADDLES	20	1.00	2.00	2.00	100.0
K1962-05	SVOC-GPC-BLANK	21	1.00	2.00	2.00	100.0
K1962-06	PEST-GPC-BLANK	22	1.00	2.00	2.00	100.0
K1962-07	PEST-GPC-BLANK-SPIKE	23	1.00	2.00	2.00	100.0
K1962-08	PCB-GPC-BLANK	24	1.00	2.00	2.00	100.0
K1962-09	PCB-GPC-BLANK-SPIKE	25	1.00	2.00	2.00	100.0
K1962-10	SVOC-GPC2-BLANK	26	1.00	2.00	2.00	100.0
K1962-11	PEST-GPC2-BLANK	27	1.00	2.00	2.00	100.0
K1962-12	PEST-GPC2-BLANK-SPIKE	28	1.00	2.00	2.00	100.0
K1962-13	PCB-GPC2-BLANK	29	1.00	2.00	2.00	100.0
K1962-14	PCB-GPC2-BLANK-SPIKE	30	1.00	2.00	2.00	100.0
K1970-01	P-1-N.WALL	31	1.14	9.65	8.8	90.0
K1970-02	P-2-E.WALL-NORTH	32	1.15	9.94	8.94	88.6
K1970-03	P-3-E.WALL-N.CENTER	33	1.16	9.84	8.77	87.7
K1970-04	P-4-E.WALL-S.CENTER	34	1.19	9.96	8.91	88.0
K1970-05	P-5-E.WALL-SOUTH	35	1.16	9.94	8.85	87.6
K1970-06	P-6-SOUTH-WALL	36	1.17	9.95	8.91	88.2
K1970-07	P-7-WEST-WALL-SOUTH	37	1.14	9.91	9.06	90.3
K1970-08	P-8-W.WALL-S.CENTER	38	1.16	9.73	8.33	83.7
K1970-09	P-9-W.WALL-N.CENTER	39	1.14	9.97	9.18	91.1



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/19/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 03/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 03/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101432

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1970-10	P-10-W.WALL-NORTH	40	1.15	9.56	8.55	88.0
K1971-01	SB-8(9-10)-031519	41	1.13	9.56	8.05	82.1
K1971-02	SB-8(14-15)-031519	42	1.19	9.72	8.25	82.8
K1971-03	SB-9(1-2)-031519	43	1.17	9.55	8.04	82.0
K1971-04	SB-9(7-8)-031519	44	1.18	9.98	8.86	87.3
K1971-05	SB-10(1-2)-031519	45	1.18	9.9	8.76	86.9
K1971-06	SB-10(6-7)-031519	46	1.15	9.94	8.54	84.1
K1971-07	SB-14(9-11)-031519	47	1.12	9.71	7.63	75.8
K1971-08	SB-14(13-15)-031519	48	1.12	9.55	3.62	29.7
K1971-09	SB-15(1-2)-031519	49	1.14	9.97	8.76	86.3
K1971-10	SB-15(6-7)-031519	50	1.18	9.98	9.00	88.9
K1971-11	SB-15(5)031519	51	1.15	9.96	8.88	87.7
K1974-01	WC4	52	1.11	9.97	9.18	91.1
K1979-01	BFD1A	53	1.1	9.94	8.68	85.7
K1980-01	WCXY2	54	1.13	9.57	8.33	85.3

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$