

**TOTAL SUSPENDED SOLIDS - SM2540D**SUPERVISOR: apatelANALYST: AparnaDate: 03/19/2019Run Number: LB101436BalanceID: WC SC-3OvenID: WC Oven-2FilterID: 15028098ThermometerID: WET-OVEN-2

TEMP1 IN:	<u>103</u>	<u>03/19/2019 07:00</u>	TEMP OUT:	<u>104</u>	<u>03/19/2019 08:05</u>
TEMP2 IN:	<u>104</u>	<u>03/19/2019 09:00</u>	TEMP OUT:	<u>104</u>	<u>03/19/2019 10:10</u>
TEMP3 IN:	<u>105</u>	<u>03/19/2019 11:00</u>	TEMP OUT:	<u>105</u>	<u>03/19/2019 12:40</u>
TEMP4 IN:	<u>105</u>	<u>03/19/2019 13:20</u>	TEMP OUT:	<u>105</u>	<u>03/19/2019 15:00</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB101436BL	LB101436BL	1.4253	1.4253	100	1.4253	1.4253	1.4253	0.0000	0
2	LB101436BS	LB101436BS	1.4012	1.4012	100	1.4520	1.4520	1.4520	0.0508	508
3	K1934-02	EFF-WW	1.3994	1.3994	300	1.3999	1.3999	1.3999	0.0005	1.7
4	K1941-01	T-1	1.3780	1.3780	400	1.3780	1.3780	1.3780	0.0000	0
5	K1941-04	T-2	1.3828	1.3828	400	1.3828	1.3828	1.3828	0.0000	0
6	K1957-01	031319-JETT-E-D-B	1.4059	1.4059	250	1.4148	1.4148	1.4148	0.0089	35.6
7	K1957-01DUP	031319-JETT-E-D-BDUP	1.3851	1.3851	250	1.3941	1.3941	1.3941	0.0090	36
8	K1957-02	031319-JETT-E-D-M	1.3968	1.3968	400	1.4076	1.4076	1.4076	0.0108	27
9	K1957-03	031319-JETT-E-D-S	1.3784	1.3784	400	1.3891	1.3891	1.3891	0.0107	26.8
10	K1957-04	031319-JETT-E-U-B	1.3971	1.3971	400	1.4128	1.4128	1.4128	0.0157	39.3
11	K1957-05	031319-JETT-E-U-M	1.3787	1.3787	400	1.3911	1.3911	1.3911	0.0124	31
12	K1957-06	031319-JETT-E-U-S	1.3755	1.3755	400	1.3835	1.3835	1.3835	0.0080	20
13	K1959-01	031519-DBRI-E-D-B	1.4048	1.4048	400	1.4243	1.4243	1.4243	0.0195	48.8
14	K1959-02	031519-DBRI-E-D-M	1.3876	1.3876	400	1.3987	1.3987	1.3987	0.0111	27.8
15	K1959-03	031519-DBRI-E-D-S	1.3854	1.3854	400	1.3988	1.3988	1.3988	0.0134	33.5



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/19/2019

Run Number: LB101436

BalanceID: WC SC-3

OvenID: WC Oven-2

FilterID: 15028098

ThermometerID: WET-OVEN-2

TEMP1 IN: 103 03/19/2019 07:00 TEMP OUT: 104 03/19/2019 08:05
TEMP2 IN: 104 03/19/2019 09:00 TEMP OUT: 104 03/19/2019 10:10
TEMP3 IN: 105 03/19/2019 11:00 TEMP OUT: 105 03/19/2019 12:40
TEMP4 IN: 105 03/19/2019 13:20 TEMP OUT: 105 03/19/2019 15:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	K1959-04	031519-DBRI-E-U-B	1.4091	1.4091	400	1.4133	1.4133	1.4133	0.0042	10.5
17	K1959-05	031519-DBRI-E-U-M	1.3766	1.3766	400	1.3850	1.3850	1.3850	0.0084	21
18	K1959-06	031519-DBRI-E-U-S	1.3899	1.3899	250	1.4022	1.4022	1.4022	0.0123	49.2
19	K1959-06DUP	031519-DBRI-E-U-SDUP	1.3861	1.3861	250	1.3984	1.3984	1.3984	0.0123	49.2
20	K1961-01	001M-WILLETS-PT-BLVD (FEB)	1.4063	1.4063	100	1.4131	1.4131	1.4131	0.0068	68

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS031919

WorkList ID : 123579

Date : 03-19-2019 07:16:12

2010/436

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/21/2019	Water	K1934-02	TSS	Cool 4 deg C	ARDM01	P51	EFF-WW	03/14/2019	SM2540 D
03/25/2019	Water	K1941-01	TSS	Cool 4 deg C	PSEG01	K31	T-1	03/18/2019	SM2540 D
03/25/2019	Water	K1941-04	TSS	Cool 4 deg C	PSEG01	K31	T-2	03/18/2019	SM2540 D
03/20/2019	Water	K1957-01	TSS	Cool 4 deg C	HDR103	F11	031319-JETT-E-D-B	03/13/2019	SM2540 D
03/20/2019	Water	K1957-02	TSS	Cool 4 deg C	HDR103	F11	031319-JETT-E-D-M	03/13/2019	SM2540 D
03/20/2019	Water	K1957-03	TSS	Cool 4 deg C	HDR103	F11	031319-JETT-E-D-S	03/13/2019	SM2540 D
03/20/2019	Water	K1957-04	TSS	Cool 4 deg C	HDR103	F11	031319-JETT-E-U-B	03/13/2019	SM2540 D
03/20/2019	Water	K1957-05	TSS	Cool 4 deg C	HDR103	F11	031319-JETT-E-U-M	03/13/2019	SM2540 D
03/20/2019	Water	K1957-06	TSS	Cool 4 deg C	HDR103	F11	031319-JETT-E-U-S	03/13/2019	SM2540 D
03/22/2019	Water	K1959-01	TSS	Cool 4 deg C	HDR103	F11	031519-DBRI-E-D-B	03/15/2019	SM2540 D
03/22/2019	Water	K1959-02	TSS	Cool 4 deg C	HDR103	F11	031519-DBRI-E-D-M	03/15/2019	SM2540 D
03/22/2019	Water	K1959-03	TSS	Cool 4 deg C	HDR103	F11	031519-DBRI-E-D-S	03/15/2019	SM2540 D
03/22/2019	Water	K1959-04	TSS	Cool 4 deg C	HDR103	F11	031519-DBRI-E-U-B	03/15/2019	SM2540 D
03/22/2019	Water	K1959-05	TSS	Cool 4 deg C	HDR103	F11	031519-DBRI-E-U-M	03/15/2019	SM2540 D
03/22/2019	Water	K1959-06	TSS	Cool 4 deg C	HDR103	F11	031519-DBRI-E-U-S	03/15/2019	SM2540 D
03/21/2019	Water	K1961-01	TSS	Cool 4 deg C	TULL01	F11	001M-WILLETS-PT-BLVD(FEI)	03/14/2019	SM2540 D

Date/Time 03/19/19, 08:00
 Received by: 
 Relinquished by: 

Date/Time 03/19/19, 11:30
 Received by: 
 Relinquished by: 