



TOTAL Dissolved Solids - SM2540C

Run Number: LB101439

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/19/2019

BalanceID: WC SC-3

Filter ID: 14466061

TEMP1 IN:	<u>103</u>	<u>03/15/2019</u>	<u>15:00</u>	TEMP OUT:	<u>103</u>	<u>03/19/2019</u>	<u>07:00</u>	OvenID1:	<u>WC Oven-2</u>	Thermo ID1:	<u>WET-OVEN#2</u>
TEMP2 IN:	<u>182</u>	<u>03/19/2019</u>	<u>07:10</u>	TEMP OUT:	<u>182</u>	<u>03/19/2019</u>	<u>09:00</u>	OvenID2:	<u>WC-OVEN-3</u>	Thermo ID2:	<u>WET-OVEN#3</u>
TEMP3 IN:	<u>182</u>	<u>03/19/2019</u>	<u>10:30</u>	TEMP OUT:	<u>182</u>	<u>03/19/2019</u>	<u>12:05</u>	OvenID3:	<u>WC-OVEN-3</u>	Thermo ID3:	<u>WET-OVEN#3</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Weight (g)	Result mg/L
1	LB101439BL	LB101439BL	82.5635	82.5635	100	82.5635	82.5635	82.5635	0.0000	0
2	LB101439BS	LB101439BS	90.1236	90.1236	100	90.1340	90.1340	90.1340	0.0104	104
3	K1937-02	BPS1-TT-MW309I-20190313	85.1010	85.1010	100	85.1186	85.1186	85.1186	0.0176	176
4	K1937-02DUP	BPS1-TT-MW309I-20190313DUP	94.2739	94.2739	100	94.2920	94.2920	94.2920	0.0181	181
5	K1937-03	BPS1-TT-MW303D-20190313	77.6259	77.6259	100	77.6406	77.6406	77.6406	0.0147	147
6	K1937-04	BPS1-TT-MW313S-20190313	77.5708	77.5708	100	77.5823	77.5823	77.5823	0.0115	115
7	K1937-05	BPS1-TT-MW314S-20190313	90.0274	90.0274	100	90.0300	90.0300	90.0300	0.0026	26
8	K1937-08	BPS1-TT-MW310S-20190314	90.8158	90.8158	100	90.8373	90.8373	90.8373	0.0215	215
9	K1937-11	BP-MH-SW4001-SOUTH-20190314	85.4740	85.4740	100	85.4956	85.4956	85.4956	0.0216	216
10	K1937-12	BP-TT-SW4002-20190314	93.0166	93.0166	100	93.0400	93.0400	93.0400	0.0234	234
11	K1937-15	BPS1-TT-MW307S-20190314	90.0708	90.0708	100	90.0787	90.0787	90.0787	0.0079	79
12	K1937-16	BPS1-TT-MW306S-20190314	81.4928	81.4928	100	81.5027	81.5027	81.5027	0.0099	99
13	K1937-17	BPS1-TT-MW306I-20190314	83.4624	83.4624	100	83.4700	83.4700	83.4700	0.0076	76
14	K1937-18	BPS1-TT-MW202S-20190314	104.6976	104.6976	100	104.7710	104.7710	104.7710	0.0734	734
15	K1937-19	BPS1-TT-MW202I-20190314	68.8607	68.8607	100	68.8722	68.8722	68.8722	0.0115	115
16	K1937-20	BPS1-TT-MW303S-20190314	82.1392	82.1392	100	82.2219	82.2219	82.2219	0.0827	827
17	K1937-20DUP	BPS1-TT-MW303S-20190314DUP	74.4916	74.4916	100	74.5750	74.5750	74.5750	0.0834	834



TOTAL Dissolved Solids - SM2540C

Run Number: LB101439

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/19/2019

BalanceID: WC SC-3

Filter ID: 14466061

TEMP1 IN:	<u>103</u>	<u>03/15/2019</u>	<u>15:00</u>	TEMP OUT:	<u>103</u>	<u>03/19/2019</u>	<u>07:00</u>	OvenID1:	<u>WC Oven-2</u>	Thermo ID1:	<u>WET-OVEN#2</u>
TEMP2 IN:	<u>182</u>	<u>03/19/2019</u>	<u>07:10</u>	TEMP OUT:	<u>182</u>	<u>03/19/2019</u>	<u>09:00</u>	OvenID2:	<u>WC-OVEN-3</u>	Thermo ID2:	<u>WET-OVEN#3</u>
TEMP3 IN:	<u>182</u>	<u>03/19/2019</u>	<u>10:30</u>	TEMP OUT:	<u>182</u>	<u>03/19/2019</u>	<u>12:05</u>	OvenID3:	<u>WC-OVEN-3</u>	Thermo ID3:	<u>WET-OVEN#3</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = 2nd Empty Dish + Sample weight after 1.5 hr drying @180 (±2) °C (g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TDS031519-1


WorkList ID : 123494

Date : 03-15-2019 10:53:53

1310/439

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2019	Water	K1937-02	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW309I-20190313	03/13/2019	SM2540 C
03/20/2019	Water	K1937-03	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW303D-20190313	03/13/2019	SM2540 C
03/20/2019	Water	K1937-04	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW313S-20190313	03/13/2019	SM2540 C
03/20/2019	Water	K1937-05	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW314S-20190313	03/13/2019	SM2540 C
03/21/2019	Water	K1937-08	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW310S-20190314	03/14/2019	SM2540 C
03/21/2019	Water	K1937-11	TDS	Cool 4 deg C	TETR06	F11	BP-MH-SW4001-SOUTH-2019	03/14/2019	SM2540 C
03/21/2019	Water	K1937-12	TDS	Cool 4 deg C	TETR06	F11	BP-TT-SW4002-20190314	03/14/2019	SM2540 C
03/21/2019	Water	K1937-15	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW307S-20190314	03/14/2019	SM2540 C
03/21/2019	Water	K1937-16	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW306S-20190314	03/14/2019	SM2540 C
03/21/2019	Water	K1937-17	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW306I-20190314	03/14/2019	SM2540 C
03/21/2019	Water	K1937-18	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW202S-20190314	03/14/2019	SM2540 C
03/21/2019	Water	K1937-19	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW202I-20190314	03/14/2019	SM2540 C
03/21/2019	Water	K1937-20	TDS	Cool 4 deg C	TETR06	F11	BPS1-TT-MW303S-20190314	03/14/2019	SM2540 C

Date/Time 03/15/19 14:20
 Received by: 
 Relinquished by: 

Date/Time 03/15/19 15:30
 Received by: 
 Relinquished by: 