

LB101451


7471B

INSTRUMENT ID: CUI

Linear

 μ Abs.:

59244



A= 0.0000e+000

B= 1.6890e-004

C= 1.2886e-002

Rho= 0.9995912

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	-7	0.000	-7				
0.2	0.200	0.218	0.018	1217	0.0 %	1217				
2.5	2.500	2.591	0.091	15263	0.0 %	15263				
5.0	5.000	4.775	-0.225	28197	0.0 %	28197				
7.5	7.500	7.585	0.085	44832	0.0 %	44832				
10.0	10.000	10.019	0.019	59244	0.0 %	59244				

LB101451 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 19 Mar 2019 14:25:36				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	500	-7	-	0.0000	7471B	PPB	19 Mar 2019 15:05:45	S Std
0.2 - 1	50.2	1217	-	0.2000	7471B	PPB	19 Mar 2019 15:08:02	S Std
2.5 - 1	52.5	15263	-	2.5000	7471B	PPB	19 Mar 2019 15:10:19	S Std
5.0 - 1	55.0	28197	-	5.0000	7471B	PPB	19 Mar 2019 15:12:36	S Std
7.5 - 1	57.5	44832	-	7.5000	7471B	PPB	19 Mar 2019 15:14:53	S Std
10.0 - 1	510.0	59244	-	10.0000	7471B	PPB	19 Mar 2019 15:21:42	S Std
ICV59 - 1	ICV59	23980	4.0630	-	7471B	PPB	19 Mar 2019 15:25:36	U SMPL
ICB59 - 1	ICB59	38	0.0193	-	7471B	PPB	19 Mar 2019 15:27:51	U SMPL
CCV94 - 1	CCV94	30768	5.2095	-	7471B	PPB	19 Mar 2019 15:30:07	U SMPL
CCB94 - 1	CCB94	-63	0.0022	-	7471B	PPB	19 Mar 2019 15:32:22	U SMPL
CRA - 1	CRA	1202	0.2159	-	7471B	PPB	19 Mar 2019 15:34:38	U SMPL
HighStd - 1	HighStd	53671	9.0777	-	7471B	PPB	19 Mar 2019 15:41:56	U SMPL
ChkStd - 1	ChkStd	42033	7.1121	-	7471B	PPB	19 Mar 2019 15:44:11	U SMPL
PB118019BL - 1	PBS	44	0.0203	-	7471B	PPB	19 Mar 2019 15:46:29	U SMPL
PB118019BS - 1	LCSS	21619	3.6643	-	7471B	PPB	19 Mar 2019 15:48:46	U SMPL
K1935-01 - 1	SB-17(0.5-1.5)	11856	2.0153	-	7471B	PPB	19 Mar 2019 15:51:02	U SMPL
K1935-01DUP - 1	SB-17(0.5-1.5)DUP	12124	2.0606	-	7471B	PPB	19 Mar 2019 15:53:19	U SMPL
K1935-01MS - 1	SB-17(0.5-1.5)MS	31059	5.2586	-	7471B	PPB	19 Mar 2019 15:55:36	U SMPL
K1935-01MSD - 1	SB-17(0.5-1.5)MSD	31365	5.3103	-	7471B	PPB	19 Mar 2019 15:57:54	U SMPL
K1935-02 - 1	SB-17(5.5-6.5)	25897	4.3868	-	7471B	PPB	19 Mar 2019 16:00:10	U SMPL
CCV95 - 1	CCV95	30701	5.1982	-	7471B	PPB	19 Mar 2019 16:02:26	U SMPL
CCB95 - 1	CCB95	-90	-0.0023	-	7471B	PPB	19 Mar 2019 16:04:42	U SMPL
K1935-03 - 1	SB-5(7-8)	97771	16.5261	-	7471B	PPB	19 Mar 2019 16:06:59	U SMPL
K1935-04 - 1	SB-5(13-14)	556	0.1068	-	7471B	PPB	19 Mar 2019 16:09:14	U SMPL
K1935-06 - 1	SB-18(11-12)	7432	1.2681	-	7471B	PPB	19 Mar 2019 16:11:32	U SMPL
K1935-07 - 1	SB-18(14-15)	1609	0.2846	-	7471B	PPB	19 Mar 2019 16:13:48	U SMPL
K1935-08 - 1	SB-2(10-11)	3924	0.6756	-	7471B	PPB	19 Mar 2019 16:16:04	U SMPL
K1935-09 - 1	SB-2(14-15)	1389	0.2475	-	7471B	PPB	19 Mar 2019 16:18:21	U SMPL
K1935-10 - 1	SB-7(6-7)	167766	28.3480	-	7471B	PPB	19 Mar 2019 16:20:38	U SMPL
K1935-11 - 1	SB-7(9-10)	1095	0.1978	-	7471B	PPB	19 Mar 2019 16:22:55	U SMPL
K1951-01 - 1	FK-GF-03-006-A	5408	0.9263	-	7471B	PPB	19 Mar 2019 16:25:14	U SMPL
K1951-03 - 1	FK-GF-03-006-B	6153	1.0521	-	7471B	PPB	19 Mar 2019 16:27:30	U SMPL
CCV96 - 1	CCV96	30726	5.2024	-	7471B	PPB	19 Mar 2019 16:29:46	U SMPL
CCB96 - 1	CCB96	5	0.0137	-	7471B	PPB	19 Mar 2019 16:32:01	U SMPL
K1956-01 - 1	TP-7	1061	0.1921	-	7471B	PPB	19 Mar 2019 16:34:17	U SMPL
K1956-05 - 1	TP-8	467	0.0918	-	7471B	PPB	19 Mar 2019 16:36:33	U SMPL
K1958-01 - 1	TP-1	2115	0.3701	-	7471B	PPB	19 Mar 2019 16:38:49	U SMPL
K1958-03 - 1	TP-2	1191	0.2140	-	7471B	PPB	19 Mar 2019 16:41:04	U SMPL
K1958-05 - 1	TP-3	539	0.1039	-	7471B	PPB	19 Mar 2019 16:43:20	U SMPL
PB118020BL - 1	PBS	175	0.0424	-	7471B	PPB	19 Mar 2019 16:45:37	U SMPL
PB118020BS - 1	LCSS	21838	3.7012	-	7471B	PPB	19 Mar 2019 16:47:53	U SMPL
K1687-07 - 1	HR-05-031519-A	6257	1.0697	-	7471B	PPB	19 Mar 2019 16:50:10	U SMPL
K1687-07DUP - 1	HR-05-031519-ADUP	6352	1.0857	-	7471B	PPB	19 Mar 2019 16:52:28	U SMPL
K1687-07MS - 1	HR-05-031519-AMS	27950	4.7335	-	7471B	PPB	19 Mar 2019 16:54:44	U SMPL
CCV97 - 1	CCV97	28835	4.8830	-	7471B	PPB	19 Mar 2019 16:56:59	U SMPL
CCB97 - 1	CCB97	-33	0.0073	-	7471B	PPB	19 Mar 2019 16:59:15	U SMPL
K1687-07MSD - 1	HR-05-031519-AMSD	28491	4.8249	-	7471B	PPB	19 Mar 2019 17:01:32	U SMPL
K1687-09 - 1	HR-05-031519-B	6461	1.1041	-	7471B	PPB	19 Mar 2019 17:03:48	U SMPL
K1687-11 - 1	HR-05-031519-C	6110	1.0448	-	7471B	PPB	19 Mar 2019 17:06:04	U SMPL
K1691-09 - 1	CL-01-031519-A	6406	1.0948	-	7471B	PPB	19 Mar 2019 17:08:20	U SMPL
K1691-11 - 1	CL-01-031519-B	23027	3.9021	-	7471B	PPB	19 Mar 2019 17:10:36	U SMPL
K1691-13 - 1	CL-01-031519-C	24724	4.1887	-	7471B	PPB	19 Mar 2019 17:12:52	U SMPL
K1691-15 - 1	CL-01-031519-D	5961	1.0197	-	7471B	PPB	19 Mar 2019 17:15:09	U SMPL
K1691-17 - 1	CL-01-031519-E	11525	1.9594	-	7471B	PPB	19 Mar 2019 17:17:26	U SMPL
K1697-09 - 1	SU-01-031519-A	7595	1.2957	-	7471B	PPB	19 Mar 2019 17:19:43	U SMPL
K1697-11 - 1	SU-01-031519-B	1934	0.3395	-	7471B	PPB	19 Mar 2019 17:21:59	U SMPL
CCV98 - 1	CCV98	29786	5.0436	-	7471B	PPB	19 Mar 2019 17:24:15	U SMPL
CCB98 - 1	CCB98	-42	0.0058	-	7471B	PPB	19 Mar 2019 17:26:31	U SMPL
K1931-01 - 1	LSP-1	2238	0.3909	-	7471B	PPB	19 Mar 2019 17:28:47	U SMPL
K1933-01 - 1	S-1	3310	0.5719	-	7471B	PPB	19 Mar 2019 17:31:03	U SMPL
K1933-02 - 1	S-2	4949	0.8488	-	7471B	PPB	19 Mar 2019 17:33:19	U SMPL
K1933-03 - 1	S-3	6174	1.0557	-	7471B	PPB	19 Mar 2019 17:35:35	U SMPL
K1933-04 - 1	S-4	3363	0.5809	-	7471B	PPB	19 Mar 2019 17:37:50	U SMPL
K1933-05 - 1	S-5	7857	1.3399	-	7471B	PPB	19 Mar 2019 17:40:07	U SMPL
K1933-06 - 1	S-6	12393	2.1060	-	7471B	PPB	19 Mar 2019 17:42:23	U SMPL
K1933-07 - 1	S-7	7095	1.2112	-	7471B	PPB	19 Mar 2019 17:44:41	U SMPL
K1933-08 - 1	S-8	22941	3.8875	-	7471B	PPB	19 Mar 2019 17:46:58	U SMPL
K1933-09 - 1	S-9	1207	0.2167	-	7471B	PPB	19 Mar 2019 17:49:14	U SMPL
CCV99 - 1	CCV99	30300	5.1305	-	7471B	PPB	19 Mar 2019 17:54:48	U SMPL
CCB99 - 1	CCB99	-30	0.0078	-	7471B	PPB	19 Mar 2019 17:57:04	U SMPL
K1935-01LX5 - 1		2350	0.4098	-	7471B	PPB	19 Mar 2019 17:59:21	U SMPL
K1935-01A - 1		31735	5.3728	-	7471B	PPB	19 Mar 2019 18:01:37	U SMPL
K1687-07LX5 - 1		1165	0.2097	-	7471B	PPB	19 Mar 2019 18:03:53	U SMPL
K1687-07A - 1		28211	4.7776	-	7471B	PPB	19 Mar 2019 18:06:11	U SMPL
K1935-03DLX2 - 1		48972	8.2841	-	7471B	PPB	19 Mar 2019 18:08:27	U SMPL
K1935-10DLX5 - 1		34962	5.9178	-	7471B	PPB	19 Mar 2019 18:10:44	U SMPL
CCV01 - 1	CCV01	29866	5.0571	-	7471B	PPB	19 Mar 2019 18:13:02	U SMPL
CCB01 - 1	CCB01	-77	-0.0001	-	7471B	PPB	19 Mar 2019 18:15:20	U SMPL