



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101455

Ambient Barometric Pressure (mmHg): 765

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.3	03/19/2019	10:05
2	K1941-02		1	9	>100.0	>212.0	>212.0	03/19/2019	10:40
3	K1941-02DUP		1	9	>100.0	>212.0	>212.0	03/19/2019	11:05
4	K1941-05		1	7	>100.0	>212.0	>212.0	03/19/2019	11:32

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB101455

WorkList Name : FP31919

WorkList ID : 123623

Date : 03-19-2019 09:27:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/25/2019	Water	K1941-02	C Flash Point	Cool 4 deg C	PSEG01	K31	T-1	03/18/2019	1010A
03/25/2019	Water	K1941-05	B Flash Point	Cool 4 deg C	PSEG01	K31	T-2	03/18/2019	1010A

Date/Time 03/19/19 09:30
 Received by: KS
 Relinquished by: CP

Date/Time 03/19/19 11:40
 Received by: CP
 Relinquished by: KS