

LB101469


245.1

INSTRUMENT ID: C01

Linear

 μ Abs.:

41067



A= 0.0000e+000

B= 2.4576e-004

C= -1.6690e-002

Rho= 0.9993504

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	8	0.000	8				
0.05	0.050	0.043	-0.007	242	0.0 %	242				
0.2	0.200	0.190	-0.010	842	0.0 %	842				
2.5	2.500	2.499	-0.001	10238	0.0 %	10238				
5.0	5.000	5.221	0.221	21314	0.0 %	21314				
7.5	7.500	7.235	-0.265	29508	0.0 %	29508				
10.0	10.000	10.076	0.076	41067	0.0 %	41067				

LB101469 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Mar 2019 10:23:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	8	-	0.0000	245.1	PPB	20 Mar 2019 10:27:02	S	Std
0.05 - 1	50.05	242	-	0.0500	245.1	PPB	20 Mar 2019 10:29:19	S	Std
0.2 - 1	50.2	842	-	0.2000	245.1	PPB	20 Mar 2019 10:31:36	S	Std
2.5 - 1	52.5	10238	-	2.5000	245.1	PPB	20 Mar 2019 10:33:53	S	Std
5.0 - 1	55.0	21314	-	5.0000	245.1	PPB	20 Mar 2019 10:36:10	S	Std
7.5 - 1	57.5	29508	-	7.5000	245.1	PPB	20 Mar 2019 10:38:26	S	Std
10.0 - 1	510.0	41067	-	10.0000	245.1	PPB	20 Mar 2019 10:48:13	S	Std
ICV60 - 1	ICV60	15801	3.8665	-245.1		PPB	20 Mar 2019 10:53:07	U	SMPL
ICB60 - 1	ICB60	-10	-0.0191	-245.1		PPB	20 Mar 2019 10:55:22	U	SMPL
CCV02 - 1	CCV02	20700	5.0705	-245.1		PPB	20 Mar 2019 10:57:37	U	SMPL
CCB02 - 1	CCB02	-49	-0.0287	-245.1		PPB	20 Mar 2019 10:59:51	U	SMPL
CRA - 1	CRA	240	0.0423	-245.1		PPB	20 Mar 2019 11:06:47	U	SMPL
HighStd - 1	HighStd	40168	9.8549	-245.1		PPB	20 Mar 2019 11:09:01	U	SMPL
ChkStd - 1	ChkStd	28034	6.8729	-245.1		PPB	20 Mar 2019 11:11:17	U	SMPL
PB118022BL - 1	PBW	-31	-0.0243	-245.1		PPB	20 Mar 2019 11:13:34	U	SMPL
PB118022BS - 1	LCSW	16763	4.1029	-245.1		PPB	20 Mar 2019 11:15:52	U	SMPL
K1903-01 - 1	TW-1	131	0.0155	-245.1		PPB	20 Mar 2019 11:18:09	U	SMPL
K1903-01DUP - 1	TW-1DUP	190	0.0300	-245.1		PPB	20 Mar 2019 11:20:25	U	SMPL
K1903-01MS - 1	TW-1MS	17483	4.2799	-245.1		PPB	20 Mar 2019 11:22:43	U	SMPL
K1903-01MSD - 1	TW-1MSD	17253	4.2234	-245.1		PPB	20 Mar 2019 11:24:59	U	SMPL
K1903-02 - 1	TW-3	57	-0.0027	-245.1		PPB	20 Mar 2019 11:27:15	U	SMPL
CCV03 - 1	CCV03	21022	5.1496	-245.1		PPB	20 Mar 2019 11:29:30	U	SMPL
CCB03 - 1	CCB03	-33	-0.0248	-245.1		PPB	20 Mar 2019 11:31:46	U	SMPL
K1903-03 - 1	TW-6	25045	6.1383	-245.1		PPB	20 Mar 2019 11:34:01	U	SMPL
K1903-05 - 1	TW-1	16	-0.0128	-245.1		PPB	20 Mar 2019 11:36:16	U	SMPL
K1903-06 - 1	TW-3	102	0.0084	-245.1		PPB	20 Mar 2019 11:38:31	U	SMPL
K1903-07 - 1	TW-6	86	0.0044	-245.1		PPB	20 Mar 2019 11:40:47	U	SMPL
K1909-01 - 1	MH-02	148	0.0197	-245.1		PPB	20 Mar 2019 11:43:03	U	SMPL
K1909-02 - 1	101-102	106	0.0094	-245.1		PPB	20 Mar 2019 11:45:19	U	SMPL
K1909-03 - 1	93-94	97	0.0071	-245.1		PPB	20 Mar 2019 11:47:36	U	SMPL
K1909-04 - 1	97-98	109	0.0101	-245.1		PPB	20 Mar 2019 11:49:53	U	SMPL
K1909-05 - 1	MH-CO5(SETTLE-WATER)	130	0.0153	-245.1		PPB	20 Mar 2019 11:52:10	U	SMPL
K1932-01 - 1	TW-2	712	0.1583	-245.1		PPB	20 Mar 2019 11:54:26	U	SMPL
CCV04 - 1	CCV04	20988	5.1413	-245.1		PPB	20 Mar 2019 11:56:42	U	SMPL
CCB04 - 1	CCB04	-17	-0.0209	-245.1		PPB	20 Mar 2019 11:58:58	U	SMPL
K1932-02 - 1	TW-7	607	0.1325	-245.1		PPB	20 Mar 2019 12:01:13	U	SMPL
K1932-04 - 1	TW-5	211	0.0352	-245.1		PPB	20 Mar 2019 12:03:29	U	SMPL
K1932-05 - 1	TW-2	166	0.0241	-245.1		PPB	20 Mar 2019 12:05:45	U	SMPL
K1932-06 - 1	TW-7	64	-0.0010	-245.1		PPB	20 Mar 2019 12:08:00	U	SMPL
K1932-07 - 1	TW-5	77	0.0022	-245.1		PPB	20 Mar 2019 12:10:17	U	SMPL
K1903-01LX5 - 1		26	-0.0103	-245.1		PPB	20 Mar 2019 12:12:34	U	SMPL
K1903-01A - 1		17232	4.2182	-245.1		PPB	20 Mar 2019 12:14:50	U	SMPL
CCV05 - 1	CCV05	21191	5.1912	-245.1		PPB	20 Mar 2019 12:17:08	U	SMPL
CCB05 - 1	CCB05	-46	-0.0280	-245.1		PPB	20 Mar 2019 12:19:25	U	SMPL