



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101476

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.0 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	03/20/2019	10:10
2	CAL2	1	Water	NA	NA	21.1	7.01	03/20/2019	10:11
3	CAL3	1	Water	NA	NA	21.1	10.02	03/20/2019	10:12
4	ICV	1	Water	NA	NA	21.1	7.00	03/20/2019	10:14
5	CCV1	1	Water	NA	NA	21.1	2.02	03/20/2019	10:15
6	K1994-04	1	Solid	20.02	20	22.3	6.43	03/20/2019	10:16
7	K1994-04DUP	1	Solid	20.03	20	22.4	6.44	03/20/2019	10:17
8	CCV2	1	Water	NA	NA	21.1	12.01	03/20/2019	10:19

LB 101476

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1994

WorkList ID : 123673

Date : 03-20-2019 09:56:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2019	Solid	K1994-04	pH	Cool 4 deg C	PSEG01	K31	50015-51054-30199-51048-51	03/19/2019	9045D

E.A

Date/Time 03/20/19 10:00
Received by: Jn
Relinquished by: Cd

Date/Time 03/20/19 11:10
Received by: Jp
Relinquished by: Cd