

TOTAL VOLATILE SOLIDS - 160.4

TEMP1 IN: 103 03/20/2019 15:32 **TEMP OUT:** 105 03/21/2019 07:10
TEMP2 IN: 105 03/21/2019 08:30 **TEMP OUT:** 105 03/21/2019 10:00
TEMP3 IN: 550 03/21/2019 11:30 **TEMP OUT:** 105 03/21/2019 15:00
TEMP4 IN: 550 03/21/2019 16:40 **TEMP OUT:** 550 03/21/2019 20:00

Run Number: LB101488

SUPERVISOR: apatel

ANALYST: Aparna

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Empty Dish + Sample Weight (g)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result %
1	LB101488BL	60.2536	60.2536	60.2536	60.2536	60.2536	60.2536	60.2536	0	0
2	K1943-01	59.9829	59.9829	84.9674	65.7205	65.7205	64.2134	64.2134	1.5071	26.3
3	K1943-02	59.5627	59.5627	84.0621	66.0555	66.0555	64.5217	64.5217	1.5338	23.6
4	K1943-03	59.3975	59.3975	83.745	67.5503	67.5503	66.2263	66.2263	1.324	16.2
5	K1943-04	53.5533	53.5533	77.548	77.548	58.0233	58.0233	56.9915	1.0318	23.1
6	K1943-05	57.1845	57.1845	81.5155	61.0553	61.0553	59.5002	59.5002	1.5551	40.2
7	K1943-06	68.0029	68.0029	84.7071	70.6394	70.6394	69.8748	69.8748	0.7646	29
8	K1943-07	60.8285	60.8285	83.352	65.6035	65.6035	64.7859	64.7859	0.8176	17.1
9	K1943-08	60.8916	60.8916	83.6897	68.6499	68.6499	67.6298	67.6298	1.0201	13.1
10	K1943-09	60.9431	60.9431	83.9829	67.2122	67.2122	65.9194	65.9194	1.2928	20.6
11	K1943-09DUP	55.8878	55.8878	79.9414	62.2952	62.2952	60.9618	60.9618	1.3334	20.8

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result mg/L = $\frac{D}{F - E} \star 100$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS032019

WorkList ID : 123651

Date : 03-20-2019 07:50:44

13101488

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/22/2019	Solid	K1943-01	TVS	Cool 4 deg C	AQUA04	F22	20190124-LOC-1	03/11/2019	160.4
03/22/2019	Solid	K1943-02	TVS	Cool 4 deg C	AQUA04	F22	20190125-LOC-2	03/11/2019	160.4
03/22/2019	Solid	K1943-03	TVS	Cool 4 deg C	AQUA04	F22	20190126-LOC-3	03/11/2019	160.4
03/22/2019	Solid	K1943-04	TVS	Cool 4 deg C	AQUA04	F22	20190127-LOC-4	03/11/2019	160.4
03/22/2019	Solid	K1943-05	TVS	Cool 4 deg C	AQUA04	F22	20190128-LOC-5	03/12/2019	160.4
03/22/2019	Solid	K1943-06	TVS	Cool 4 deg C	AQUA04	F22	20190129-LOC-6	03/12/2019	160.4
03/22/2019	Solid	K1943-07	TVS	Cool 4 deg C	AQUA04	F22	20190130-LOC-7	03/12/2019	160.4
03/22/2019	Solid	K1943-08	TVS	Cool 4 deg C	AQUA04	F22	20190131-LOC-8	03/12/2019	160.4
03/22/2019	Solid	K1943-09	TVS	Cool 4 deg C	AQUA04	F22	20190132-LOC-9	03/13/2019	160.4

Date/Time

Received by:

Relinquished by:

03/20/19 14:30
AS
AS

Date/Time

Received by:

Relinquished by:

03/20/19 16:00
AS
AS