



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101496

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/20/2019	14:20
2	CAL2	1	Water	NA	NA	20.2	7.01	03/20/2019	14:21
3	CAL3	1	Water	NA	NA	20.2	10.02	03/20/2019	14:22
4	ICV	1	Water	NA	NA	20.2	6.99	03/20/2019	17:24
5	CCV1	1	Water	NA	NA	20.2	2.01	03/20/2019	14:25
6	K2026-01	1	Water	NA	NA	19.8	4.37	03/20/2019	14:26
7	K2026-01DUP	1	Water	NA	NA	19.9	4.38	03/20/2019	14:27
8	K2026-02	1	Water	NA	NA	20.2	4.61	03/20/2019	14:30
9	K2026-03	1	Water	NA	NA	20.5	4.74	03/20/2019	14:31
10	CCV2	1	Water	NA	NA	20.2	12.01	03/20/2019	14:32

2-13 101496

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2026

WorkList ID : 123690

Date : 03-20-2019 14:03:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2019	Water	K2026-01-17	pH	Cool 4 deg C	STER07	F41	CELL-A	03/19/2019	SM 4500-H B
03/19/2019	Water	K2026-02-17	pH	Cool 4 deg C	STER07	F41	CELL-B/C	03/19/2019	SM 4500-H B
03/19/2019	Water	K2026-03-17	pH	Cool 4 deg C	STER07	F41	CELL-D	03/19/2019	SM 4500-H B

Date/Time 03-20-19 14:10
Received by: JB
Relinquished by: CB

Date/Time 03-20-19 15:10
Received by: JB
Relinquished by: CB