



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/21/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:25
In Date: 03/20/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 03/21/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101497

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1690-01	HD-02-032019	1	1.12	9.82	9.02	90.8
K1694-01	EO-02-032019-A	2	1.16	9.68	8.79	89.6
K1694-03	EO-02-032019-B	3	1.14	9.97	8.93	88.2
K1694-05	EO-02-032019-C	4	1.16	9.77	8.97	90.7
K1995-01	390311071	5	1.14	9.97	1.59	5.1
K1997-01	390312022	6	1.15	9.75	1.6	5.2
K1998-01	390314723	7	1.17	9.79	3.32	24.9
K1999-01	390314023	8	1.15	9.95	1.61	5.2
K2000-01	390314722	9	1.19	9.94	1.35	1.8
K2001-01	390314028	10	1.17	9.66	2.44	15.0
K2002-01	S-1 (2-10) -COMP	11	1.12	9.87	9.43	95.0
K2002-03	S-2 (5-30) -COMP	12	1.15	9.95	9.22	91.7
K2002-05	S-1 (10)	13	1.14	9.93	9.57	95.9
K2002-07	S-2 (30)	14	1.17	9.92	8.82	87.4
K2003-01	S4A (5-30) COMP	15	1.16	9.74	9.01	91.5
K2003-10	S4A (30)	16	1.15	9.8	9.14	92.4
K2004-01	SB-20 (10-11)	17	1.17	9.9	9.09	90.7
K2004-02	SB-20 (14-15)	18	1.19	9.87	4.82	41.8
K2004-03	SB-19 (7.5-8.5)	19	1.14	9.78	8.56	85.9
K2004-04	SB-19 (13-14)	20	1.16	9.97	6.2	57.2
K2004-05	SB-24 (7-8)	21	1.14	9.78	8.05	80.0
K2004-06	SB-24 (12-13)	22	1.19	9.98	5.25	46.2
K2004-07	SB-19B (7-8)	23	1.14	9.65	8.51	86.6
K2004-08	SB-19B (13.5-14.5)	24	1.12	9.98	6.45	60.2
K2004-09	SB-22 (7-8)	25	1.19	9.65	7.54	75.1
K2004-10	SB-22 (12-13)	26	1.14	9.74	5.87	55.0
K2004-11	SB-23 (0.5-1.5)	27	1.11	9.72	8.67	87.8
K2004-12	SB-23 (4-5)	28	1.16	9.61	8.14	82.6
K2010-01	TP-1	29	1.15	9.95	8.73	86.1
K2010-02	TP-1-A	30	1.15	9.95	8.73	86.1
K2010-03	TP-1-B	31	1.15	9.95	8.73	86.1
K2010-04	TP-1	32	1.15	9.95	8.73	86.1
K2018-01	SP-1-A	33	1.17	9.57	8.59	88.3
K2018-02	SP-1-A	34	1.17	9.57	8.59	88.3
K2018-03	SP-1-B	35	1.13	9.73	8.44	85.0
K2018-04	SP-1-C	36	1.13	9.73	8.44	85.0
K2018-05	SP-1-C	37	1.15	9.68	8.69	88.4
K2018-06	SP-1-B	38	1.15	9.68	8.69	88.4
K2024-01	RCA-08	39	1.17	9.51	8.5	87.9



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/21/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:25
In Date: 03/20/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 03/21/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101497

<u>Lab ID</u>	<u>Client Sample ID</u>	<u>Dish#</u>	<u>Dish Wt(g)</u> <u>(A)</u>	<u>Dish +</u> <u>Sample Wt(g)</u> <u>(B)</u>	<u>Dish + Dry</u> <u>Sample Wt(g)</u> <u>(C)</u>	<u>% Solid</u>
K2024-02	RCA-09	40	1.1	9.92	8.96	89.1

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$