

LB101501


7471B

INSTRUMENT ID : CUI

Linear

 μ Abs.:

60208



A= 0.0000e+000

B= 1.6708e-004

C= -1.1496e-002

Rho= 0.9999242

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	43	0.000	43				
0.2	0.200	0.216	0.016	1364	0.0 %	1364				
2.5	2.500	2.549	0.049	15327	0.0 %	15327				
5.0	5.000	4.919	-0.081	29512	0.0 %	29512				
7.5	7.500	7.471	-0.029	44782	0.0 %	44782				
10.0	10.000	10.048	0.048	60208	0.0 %	60208				

LB101501 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Mar 2019 18:06:28

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	43	-	0.0000	7471B	PPB	20 Mar 2019 18:06:49	S	Std
0.2 - 1	50.2	1364	-	0.2000	7471B	PPB	20 Mar 2019 18:09:06	S	Std
2.5 - 1	52.5	15327	-	2.5000	7471B	PPB	20 Mar 2019 18:11:24	S	Std
5.0 - 1	55.0	29512	-	5.0000	7471B	PPB	20 Mar 2019 18:13:41	S	Std
7.5 - 1	57.5	44782	-	7.5000	7471B	PPB	20 Mar 2019 18:15:58	S	Std
10.0 - 1	510.0	60208	-	10.0000	7471B	PPB	20 Mar 2019 18:21:58	S	Std
ICV62 - 1	ICV62	24535	4.0879	-7471B	PPB	PPB	20 Mar 2019 18:26:09	U	SMPL
ICB62 - 1	ICB62	66	-0.0005	-7471B	PPB	PPB	20 Mar 2019 18:28:25	U	SMPL
CCV15 - 1	CCV15	29464	4.9114	-7471B	PPB	PPB	20 Mar 2019 18:30:41	U	SMPL
CCB15 - 1	CCB15	-22	-0.0152	-7471B	PPB	PPB	20 Mar 2019 18:32:56	U	SMPL
CRA - 1	CRA	1286	0.2034	-7471B	PPB	PPB	20 Mar 2019 18:35:12	U	SMPL
HighStd - 1	HighStd	60214	10.0492	-7471B	PPB	PPB	20 Mar 2019 18:37:27	U	SMPL
ChkStd - 1	ChkStd	41525	6.9266	-7471B	PPB	PPB	20 Mar 2019 18:39:43	U	SMPL
PB118106BL - 1	PBS	95	0.0044	-7471B	PPB	PPB	20 Mar 2019 18:42:01	U	SMPL
PB118106BS - 1	LCSS	21519	3.5840	-7471B	PPB	PPB	20 Mar 2019 18:44:18	U	SMPL
K1692-03 - 1	OK-02-031819	2207	0.3573	-7471B	PPB	PPB	20 Mar 2019 18:46:34	U	SMPL
K1692-03DUP - 1	OK-02-031819DUP	2229	0.3609	-7471B	PPB	PPB	20 Mar 2019 18:48:51	U	SMPL
K1692-03MS - 1	OK-02-031819MS	24236	4.0379	-7471B	PPB	PPB	20 Mar 2019 18:51:08	U	SMPL
K1692-03MSD - 1	OK-02-031819MSD	24294	4.0476	-7471B	PPB	PPB	20 Mar 2019 18:53:25	U	SMPL
K1693-09 - 1	OR-03-031819-A	1131	0.1775	-7471B	PPB	PPB	20 Mar 2019 18:55:41	U	SMPL
CCV16 - 1	CCV16	29200	4.8673	-7471B	PPB	PPB	20 Mar 2019 18:57:57	U	SMPL
CCB16 - 1	CCB16	-42	-0.0185	-7471B	PPB	PPB	20 Mar 2019 19:00:12	U	SMPL
K1693-11 - 1	OR-03-031819-B	1471	0.2343	-7471B	PPB	PPB	20 Mar 2019 19:02:28	U	SMPL
K1941-07 - 1	SP	597	0.0883	-7471B	PPB	PPB	20 Mar 2019 19:04:43	U	SMPL
K1974-01 - 1	WC4	5394	0.8897	-7471B	PPB	PPB	20 Mar 2019 19:06:59	U	SMPL
K1984-01 - 1	SB-12(7-8)	1062	0.1659	-7471B	PPB	PPB	20 Mar 2019 19:09:15	U	SMPL
K1984-02 - 1	SB-12(14-15)	629	0.0936	-7471B	PPB	PPB	20 Mar 2019 19:11:31	U	SMPL
K1984-03 - 1	SB-11(6-7)	612418	HIGH	-7471B	PPB	PPB	20 Mar 2019 19:13:47	X	
K1984-04 - 1	SB-11(14-15)	-154	-0.0372	-7471B	PPB	PPB	20 Mar 2019 19:15:36	U	SMPL
K1984-05 - 1	SB-13(10-11)	127971	21.3703	-7471B	PPB	PPB	20 Mar 2019 19:17:56	U	SMPL
K1984-06 - 1	SB-13(13.5-14.5)	4472	0.7357	-7471B	PPB	PPB	20 Mar 2019 19:20:14	U	SMPL
K1987-01 - 1	20190143-COMPOSITE-B	24625	4.1029	-7471B	PPB	PPB	20 Mar 2019 19:22:32	U	SMPL
CCV17 - 1	CCV17	31713	5.2872	-7471B	PPB	PPB	20 Mar 2019 19:24:47	U	SMPL
CCB17 - 1	CCB17	-65	-0.0224	-7471B	PPB	PPB	20 Mar 2019 19:27:03	U	SMPL
K1987-02 - 1	20190144-COMPOSITE-C	23203	3.8653	-7471B	PPB	PPB	20 Mar 2019 19:29:20	U	SMPL
K1987-03 - 1	20190145-COMPOSITE-D	24671	4.1106	-7471B	PPB	PPB	20 Mar 2019 19:31:36	U	SMPL
K1987-04 - 1	20190146-COMPOSITE-E	18770	3.1247	-7471B	PPB	PPB	20 Mar 2019 19:33:52	U	SMPL
K1987-05 - 1	20190147-COMPOSITE-F	22063	3.6749	-7471B	PPB	PPB	20 Mar 2019 19:36:08	U	SMPL
K1987-06 - 1	20190084-FIELD-DUPLICATE	24028	4.0032	-7471B	PPB	PPB	20 Mar 2019 19:38:24	U	SMPL
K1989-01 - 1	DRUM-1-2	120	0.0086	-7471B	PPB	PPB	20 Mar 2019 19:40:40	U	SMPL
K1994-01 - 1	30200-05	85020	14.1939	-7471B	PPB	PPB	20 Mar 2019 19:42:58	U	SMPL
K1994-02 - 1	30206-10	3978	0.6532	-7471B	PPB	PPB	20 Mar 2019 19:45:15	U	SMPL
K1994-03 - 1	30211-16	22979	3.8279	-7471B	PPB	PPB	20 Mar 2019 19:47:34	U	SMPL
K1994-05 - 1	40021	110814	18.5036	-7471B	PPB	PPB	20 Mar 2019 19:49:50	U	SMPL
CCV18 - 1	CCV18	31755	5.2942	-7471B	PPB	PPB	20 Mar 2019 19:52:07	U	SMPL
CCB18 - 1	CCB18	-122	-0.0319	-7471B	PPB	PPB	20 Mar 2019 19:54:25	U	SMPL
PB118107BL - 1	PBS	13	-0.0093	-7471B	PPB	PPB	20 Mar 2019 19:56:42	U	SMPL
PB118107BS - 1	LCSS	21740	3.6209	-7471B	PPB	PPB	20 Mar 2019 19:58:58	U	SMPL
K1688-21 - 1	JC-02-031919-A	5575	0.9200	-7471B	PPB	PPB	20 Mar 2019 20:01:13	U	SMPL
K1688-21DUP - 1	JC-02-031919-ADUP	5556	0.9168	-7471B	PPB	PPB	20 Mar 2019 20:03:30	U	SMPL
K1688-21MS - 1	JC-02-031919-AMS	29641	4.9410	-7471B	PPB	PPB	20 Mar 2019 20:05:46	U	SMPL
K1688-21MSD - 1	JC-02-031919-AMSD	29706	4.9519	-7471B	PPB	PPB	20 Mar 2019 20:08:03	U	SMPL
K1688-23 - 1	JC-02-031919-B	7483	1.2388	-7471B	PPB	PPB	20 Mar 2019 20:10:20	U	SMPL
K1688-25 - 1	JC-02-031919-C	9187	1.5235	-7471B	PPB	PPB	20 Mar 2019 20:12:38	U	SMPL
K1688-27 - 1	JC-02-031919-D	18388	3.0608	-7471B	PPB	PPB	20 Mar 2019 20:14:55	U	SMPL
K1688-29 - 1	JC-02-031919-E	13809	2.2958	-7471B	PPB	PPB	20 Mar 2019 20:17:11	U	SMPL
CCV19 - 1	CCV19	30781	5.1315	-7471B	PPB	PPB	20 Mar 2019 20:19:27	U	SMPL
CCB19 - 1	CCB19	-63	-0.0220	-7471B	PPB	PPB	20 Mar 2019 20:21:44	U	SMPL
K1688-31 - 1	JC-02-031919-F	12828	2.1318	-7471B	PPB	PPB	20 Mar 2019 20:24:02	U	SMPL
K1688-33 - 1	JC-02-031919-G	9452	1.5678	-7471B	PPB	PPB	20 Mar 2019 20:26:24	U	SMPL
K1688-35 - 1	JC-02-031919-H	17729	2.9507	-7471B	PPB	PPB	20 Mar 2019 20:28:40	U	SMPL
K1688-37 - 1	JC-02-031919-I	6008	0.9923	-7471B	PPB	PPB	20 Mar 2019 20:30:57	U	SMPL
K1688-39 - 1	JC-02-031919-J	11668	1.9380	-7471B	PPB	PPB	20 Mar 2019 20:33:15	U	SMPL
K1971-01 - 1	SB-8(9-10)-031519	13234	2.1997	-7471B	PPB	PPB	20 Mar 2019 20:35:33	U	SMPL
K1971-02 - 1	SB-8(14-15)-031519	3773	0.6189	-7471B	PPB	PPB	20 Mar 2019 20:37:51	U	SMPL
K1971-03 - 1	SB-9(1-2)-031519	2537	0.4124	-7471B	PPB	PPB	20 Mar 2019 20:40:09	U	SMPL
K1971-04 - 1	SB-9(7-8)-031519	23270	3.8765	-7471B	PPB	PPB	20 Mar 2019 20:42:28	U	SMPL
K1971-05 - 1	SB-10(1-2)-031519	18306	3.0471	-7471B	PPB	PPB	20 Mar 2019 20:44:45	U	SMPL
CCV20 - 1	CCV20	29540	4.9241	-7471B	PPB	PPB	20 Mar 2019 20:47:02	U	SMPL
CCB20 - 1	CCB20	-52	-0.0202	-7471B	PPB	PPB	20 Mar 2019 20:49:19	U	SMPL
K1971-06 - 1	SB-10(6-7)-031519	43246	7.2142	-7471B	PPB	PPB	20 Mar 2019 20:51:37	U	SMPL
K1971-07 - 1	SB-14(9-11)-031519	50053	8.3515	-7471B	PPB	PPB	20 Mar 2019 20:53:54	U	SMPL
K1971-08 - 1	SB-14(13-15)-031519	999	0.1554	-7471B	PPB	PPB	20 Mar 2019 20:56:12	U	SMPL
K1971-09 - 1	SB-15(1-2)-031519	51337	8.5660	-7471B	PPB	PPB	20 Mar 2019 20:58:31	U	SMPL
K1971-10 - 1	SB-15(6-7)-031519	34724	5.7903	-7471B	PPB	PPB	20 Mar 2019 21:00:48	U	SMPL
K1692-03LX5 - 1		315	0.0411	-7471B	PPB	PPB	20 Mar 2019 21:03:06	U	SMPL
K1692-03A - 1		23767	3.9596	-7471B	PPB	PPB	20 Mar 2019 21:05:24	U	SMPL
K1688-21LX5 - 1		1013	0.1578	-7471B	PPB	PPB	20 Mar 2019 21:07:41	U	SMPL
K1688-21A - 1		28855	4.8097	-7471B	PPB	PPB	20 Mar 2019 21:09:59	U	SMPL
K1984-03DLX10 - 1		108998	18.2002	-7471B	PPB	PPB	20 Mar 2019 21:12:16	U	SMPL
CCV21 - 1	CCV21	30690	5.1163	-7471B	PPB	PPB	20 Mar 2019 21:14:34	U	SMPL
CCB21 - 1	CCB21	-81	-0.0250	-7471B	PPB	PPB	20 Mar 2019 21:16:54	U	SMPL
K1984-03DL2X50 - 1		22250	3.7061	-7471B	PPB	PPB	20 Mar 2019 21:19:11	U	SMPL
K1984-05DLX5 - 1		25035	4.1714	-7471B	PPB	PPB	20 Mar 2019 21:21:27	U	SMPL
K1994-01DLX2 - 1		44158	7.3666	-7471B	PPB	PPB	20 Mar 2019 21:23:44	U	SMPL
K1994-05DLX6 - 1		18091	3.0112	-7471B	PPB	PPB	20 Mar 2019 21:26:02	U	SMPL
CCV22 - 1	CCV22	29827	4.9721	-7471B	PPB	PPB	20 Mar 2019 21:28:21	U	SMPL
CCB22 - 1	CCB22	-51	-0.0200	-7471B	PPB	PPB	20 Mar 2019 21:30:37	U	SMPL