

LB 10/510


7470A

INSTRUMENT ID: CU1

Linear

 μ Abs.:

42172



A= 0.0000e+000

B= 2.3967e-004

C= 2.4470e-003

Rho= 0.9997249

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	28	0.000	28				
0.2	0.200	0.216	0.016	889	0.0 %	889				
2.5	2.500	2.512	0.012	10470	0.0 %	10470				
5.0	5.000	5.031	0.031	20983	0.0 %	20983				
7.5	7.500	7.323	-0.177	30543	0.0 %	30543				
10.0	10.000	10.110	0.110	42172	0.0 %	42172				

LB101510 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Mar 2019 10:41:59

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	28	-	0.0000	7470A	PPB	21 Mar 2019 12:30:37	S	Std
0.2 - 1	50.2	889	-	0.2000	7470A	PPB	21 Mar 2019 12:32:45	S	Std
2.5 - 1	52.5	10470	-	2.5000	7470A	PPB	21 Mar 2019 12:34:53	S	Std
5.0 - 1	55.0	20983	-	5.0000	7470A	PPB	21 Mar 2019 12:37:01	S	Std
7.5 - 1	57.5	30543	-	7.5000	7470A	PPB	21 Mar 2019 12:39:09	S	Std
10.0 - 1	510.0	42172	-	10.0000	7470A	PPB	21 Mar 2019 12:45:24	S	Std
ICV63 - 1	ICV63	17019	4.0813	-	7470A	PPB	21 Mar 2019 12:49:41	U	SMPL
ICB63 - 1	ICB63	10	0.0048	-	7470A	PPB	21 Mar 2019 12:51:47	U	SMPL
CCV23 - 1	CCV23	20694	4.9621	-	7470A	PPB	21 Mar 2019 12:53:53	U	SMPL
CCB23 - 1	CCB23	8	0.0044	-	7470A	PPB	21 Mar 2019 12:55:59	U	SMPL
CRA - 1	CRA	837	0.2030	-	7470A	PPB	21 Mar 2019 12:58:04	U	SMPL
HighStd - 1	HighStd	40557	9.7226	-	7470A	PPB	21 Mar 2019 13:00:10	U	SMPL
ChkStd - 1	ChkStd	28517	6.8370	-	7470A	PPB	21 Mar 2019 13:02:17	U	SMPL
PB118140BL - 1	PBW	-10	0.0001	-	7470A	PPB	21 Mar 2019 13:04:24	U	SMPL
PB118140BS - 1	LCSW	17044	4.0873	-	7470A	PPB	21 Mar 2019 13:06:31	U	SMPL
K1688-22 - 1	JC-02-031919-A	31	0.0099	-	7470A	PPB	21 Mar 2019 13:08:38	U	SMPL
K1688-22DUP - 1	JC-02-031919-ADUP	51	0.0147	-	7470A	PPB	21 Mar 2019 13:10:45	U	SMPL
K1688-22MS - 1	JC-02-031919-AMS	16787	4.0257	-	7470A	PPB	21 Mar 2019 13:12:53	U	SMPL
K1688-22MSD - 1	JC-02-031919-AMSD	16608	3.9828	-	7470A	PPB	21 Mar 2019 13:15:01	U	SMPL
K1688-24 - 1	JC-02-031919-B	8	0.0044	-	7470A	PPB	21 Mar 2019 13:17:07	U	SMPL
CCV24 - 1	CCV24	20434	4.8998	-	7470A	PPB	21 Mar 2019 13:19:13	U	SMPL
CCB24 - 1	CCB24	-1	0.0022	-	7470A	PPB	21 Mar 2019 13:21:19	U	SMPL
K1688-26 - 1	JC-02-031919-C	52	0.0149	-	7470A	PPB	21 Mar 2019 13:23:25	U	SMPL
K1688-28 - 1	JC-02-031919-D	41	0.0123	-	7470A	PPB	21 Mar 2019 13:25:31	U	SMPL
K1688-30 - 1	JC-02-031919-E	78	0.0211	-	7470A	PPB	21 Mar 2019 13:27:38	U	SMPL
K1688-32 - 1	JC-02-031919-F	55	0.0156	-	7470A	PPB	21 Mar 2019 13:29:44	U	SMPL
K1688-34 - 1	JC-02-031919-G	89	0.0238	-	7470A	PPB	21 Mar 2019 13:31:51	U	SMPL
K1688-36 - 1	JC-02-031919-H	64	0.0178	-	7470A	PPB	21 Mar 2019 13:33:58	U	SMPL
K1688-38 - 1	JC-02-031919-I	50	0.0144	-	7470A	PPB	21 Mar 2019 13:36:06	U	SMPL
K1688-40 - 1	JC-02-031919-J	56	0.0159	-	7470A	PPB	21 Mar 2019 13:38:14	U	SMPL
K1943-11 - 1	20190124-LOC-1	59	0.0166	-	7470A	PPB	21 Mar 2019 13:40:22	U	SMPL
K1943-12 - 1	20190125-LOC-2	70	0.0192	-	7470A	PPB	21 Mar 2019 13:42:29	U	SMPL
CCV25 - 1	CCV25	20408	4.8936	-	7470A	PPB	21 Mar 2019 13:44:35	U	SMPL
CCB25 - 1	CCB25	-4	0.0015	-	7470A	PPB	21 Mar 2019 13:46:42	U	SMPL
K1943-13 - 1	20190126-LOC-3	80	0.0216	-	7470A	PPB	21 Mar 2019 13:48:48	U	SMPL
K1943-14 - 1	20190127-LOC-4	114	0.0298	-	7470A	PPB	21 Mar 2019 13:50:55	U	SMPL
K1943-15 - 1	20190128-LOC-5	135	0.0348	-	7470A	PPB	21 Mar 2019 13:53:01	U	SMPL
K1943-16 - 1	20190129-LOC-6	156	0.0398	-	7470A	PPB	21 Mar 2019 13:55:08	U	SMPL
K1943-17 - 1	20190130-LOC-7	144	0.0370	-	7470A	PPB	21 Mar 2019 13:57:15	U	SMPL
K1943-18 - 1	20190131-LOC-8	114	0.0298	-	7470A	PPB	21 Mar 2019 13:59:22	U	SMPL
K1943-19 - 1	20190132-LOC-9	104	0.0274	-	7470A	PPB	21 Mar 2019 14:01:30	U	SMPL
K1943-20 - 1	20190133-LOC-10	52	0.0149	-	7470A	PPB	21 Mar 2019 14:03:38	U	SMPL
PB118142BL - 1	PBW	-14	-0.0009	-	7470A	PPB	21 Mar 2019 14:05:46	U	SMPL
PB118142BS - 1	LCSW	17562	4.2115	-	7470A	PPB	21 Mar 2019 14:07:53	U	SMPL
CCV26 - 1	CCV26	22007	5.2768	-	7470A	PPB	21 Mar 2019 14:09:59	U	SMPL
CCB26 - 1	CCB26	-4	0.0015	-	7470A	PPB	21 Mar 2019 14:12:06	U	SMPL
K1965-02 - 1	BPS1-TT-MW314I-20190314-F	56	0.0159	-	7470A	PPB	21 Mar 2019 14:14:13	U	SMPL
K1965-02DUP - 1	BPS1-TT-MW314I-20190314-FDUP	61	0.0171	-	7470A	PPB	21 Mar 2019 14:16:20	U	SMPL
K1965-02MS - 1	BPS1-TT-MW314I-20190314-FMS	18095	4.3392	-	7470A	PPB	21 Mar 2019 14:18:26	U	SMPL
K1965-02MSD - 1	BPS1-TT-MW314I-20190314-FMSD	18272	4.3816	-	7470A	PPB	21 Mar 2019 14:20:33	U	SMPL
K1965-07 - 1	BP-HN-MW24IR-20190315-F	-3	0.0017	-	7470A	PPB	21 Mar 2019 14:22:40	U	SMPL
K1965-09 - 1	BP-DUP05-20190315-F	47	0.0137	-	7470A	PPB	21 Mar 2019 14:24:47	U	SMPL
K1985-02 - 1	BP-EB03-20190317	48	0.0140	-	7470A	PPB	21 Mar 2019 14:26:55	U	SMPL
K1985-03 - 1	BP-EB03-20190317-F	41	0.0123	-	7470A	PPB	21 Mar 2019 14:29:03	U	SMPL
K1985-04 - 1	BPS1-TT-MW308I-20190317	83	0.0223	-	7470A	PPB	21 Mar 2019 14:31:12	U	SMPL
K1987-07 - 1	20190085-FIELD-BLANK	75	0.0204	-	7470A	PPB	21 Mar 2019 14:33:19	U	SMPL
CCV27 - 1	CCV27	22176	5.3173	-	7470A	PPB	21 Mar 2019 14:35:25	U	SMPL
CCB27 - 1	CCB27	-15	-0.0011	-	7470A	PPB	21 Mar 2019 14:37:31	U	SMPL
PB118140TB - 1	PB118076TB	53	0.0151	-	7470A	PPB	21 Mar 2019 14:39:38	U	SMPL
K1688-22LX5 - 1		-7	0.0008	-	7470A	PPB	21 Mar 2019 14:41:44	U	SMPL
K1688-22A - 1		16866	4.0447	-	7470A	PPB	21 Mar 2019 14:43:51	U	SMPL
K1965-02LX5 - 1		-19	-0.0021	-	7470A	PPB	21 Mar 2019 14:45:57	U	SMPL
K1965-02A - 1		18091	4.3382	-	7470A	PPB	21 Mar 2019 14:48:03	U	SMPL
CCV28 - 1	CCV28	20593	4.9379	-	7470A	PPB	21 Mar 2019 14:50:10	U	SMPL
CCB28 - 1	CCB28	-46	-0.0086	-	7470A	PPB	21 Mar 2019 14:52:18	U	SMPL