



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB101525

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP72773
Residual chlorine std, 0PPM	WP72769
Residual chlorine std, 0.1PPM	WP72770
Residual chlorine LOQ, 0.2PPM	WP72776
Residual chlorine std, 0.2PPM	WP72771
Residual chlorine std, and CCV , 0.4PPM	WP72772
Residual chlorine std, 1.6PPM	WP72774
Residual chlorine ICV-LCS, 0.4PPM	WP72777
Residual chlorine LOD, 0.05PPM	WP72775
Total Chlorine Powder Pillows	W2249

Intercept: 0.0000

Slope: 1.0000

Regression: 1.000000

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	03/21/2019	13:28
2	CAL2	0.1	1	0.000	0.100	0.100	03/21/2019	13:30
3	CAL3	0.2	1	0.000	0.200	0.200	03/21/2019	13:32
4	CAL4	0.4	1	0.000	0.400	0.400	03/21/2019	13:34
5	CAL5	0.8	1	0.000	0.800	0.800	03/21/2019	13:36
6	CAL6	1.6	1	0.000	1.600	1.600	03/21/2019	13:38



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB101525

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.400	0.400	0.400	03/21/2019	13:40
2	ICB				1	0.000	0.000	0.000	0.000	03/21/2019	13:43
3	CCV1			0.4	1	0.000	0.400	0.400	0.400	03/21/2019	13:46
4	CCB1				1	0.000	0.000	0.000	0.000	03/21/2019	13:48
5	LB101525BL	50	50		1	0.000	0.000	0.000	0.000	03/21/2019	13:50
6	LB101525BS	50	50	0.4	1	0.000	0.400	0.400	0.400	03/21/2019	13:52
7	K1560-07	50	50		1	0.000	0.050	0.050	0.050	03/21/2019	13:54
8	K1560-08	50	50		1	0.000	0.210	0.210	0.210	03/21/2019	13:56
9	K1833-19	50	50		1	0.000	1.160	1.160	1.160	03/21/2019	14:00
10	CCV2			0.4	1	0.000	0.400	0.400	0.400	03/21/2019	14:03
11	CCB2				1	0.000	0.000	0.000	0.000	03/21/2019	14:05