

OVEN TEMP1: 103 °C TEMP: 105 °C

DATE & TIME IN: 03/21/2019 15:30 OUT: 03/22/2019 07:30

OVEN TEMP2: 105 °C TEMP: 105 °C

DATE & TIME IN: 03/22/2019 08:40 OUT: 03/22/2019 10:20

Run Number: LB101527

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB101527BL	LB101527BL	90.6532	90.6532	100	90.6532	90.6532	90.6532	0.0000	0
2	K2038-01	37-ST-2019	78.0780	78.0780	100	78.2377	78.2377	78.2377	0.1597	1597
3	K2038-01DUP	37-ST-2019DUP	81.4867	81.4867	100	81.6395	81.6395	81.6395	0.1528	1528

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts032119-1

WorkList ID : 123744

Date : 03-21-2019 09:36:21

20101527

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/27/2019	Water	K2038-01	TS	Cool 4 deg C	DIST01		37-ST-2019	03/20/2019	SM2540 B

Date/Time 03/21/19 15:00
 Received by: 
 Relinquished by: _____

Date/Time 03/21/19 16:00
 Received by: 
 Relinquished by: _____