

LB101532


245.1

INSTRUMENT ID : CU1

Linear

 μ Abs.:

4076E



A= 0.0000e+000

B= 2.4583e-004

C= -5.4459e-002

Rho= 0.9998211

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.053	-0.053	5	0.000	5				
0.05	0.050	0.036	-0.014	366	0.0 %	366				
0.2	0.200	0.186	-0.014	977	0.0 %	977				
2.5	2.500	2.536	0.036	10539	0.0 %	10539				
5.0	5.000	5.153	0.153	21182	0.0 %	21182				
7.5	7.500	7.426	-0.074	30428	0.0 %	30428				
10.0	10.000	9.967	-0.033	40766	0.0 %	40766				

LB101532 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 21 Mar 2019 15:21:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	5	-	0.0000	245.1	PPB	21 Mar 2019 16:04:08	S	Std
0.05 - 1	50.05	366	-	0.0500	245.1	PPB	21 Mar 2019 16:06:26	S	Std
0.2 - 1	50.2	977	-	0.2000	245.1	PPB	21 Mar 2019 16:08:43	S	Std
2.5 - 1	52.5	10539	-	2.5000	245.1	PPB	21 Mar 2019 16:11:00	S	Std
5.0 - 1	55.0	21182	-	5.0000	245.1	PPB	21 Mar 2019 16:13:17	S	Std
7.5 - 1	57.5	30428	-	7.5000	245.1	PPB	21 Mar 2019 16:15:33	S	Std
10.0 - 1	510.0	40766	-	10.0000	245.1	PPB	21 Mar 2019 16:31:16	S	Std
ICV64 - 1	ICV64	16130	3.9108	-245.1		PPB	21 Mar 2019 16:36:01	U	SMPL
ICB64 - 1	ICB64	109	-0.0277	-245.1		PPB	21 Mar 2019 16:38:17	U	SMPL
CCV29 - 1	CCV29	21097	5.1319	-245.1		PPB	21 Mar 2019 16:40:31	U	SMPL
CCB29 - 1	CCB29	164	-0.0141	-245.1		PPB	21 Mar 2019 16:42:47	U	SMPL
CRA - 1	CRA	376	0.0380	-245.1		PPB	21 Mar 2019 16:46:28	U	SMPL
HighStd - 1	HighStd	40645	9.9374	-245.1		PPB	21 Mar 2019 16:53:04	U	SMPL
ChkStd - 1	ChkStd	28470	6.9444	-245.1		PPB	21 Mar 2019 16:55:19	U	SMPL
PB118160BL - 1	PBW	77	-0.0355	-245.1		PPB	21 Mar 2019 16:57:36	U	SMPL
PB118160BS - 1	LCSW	17478	4.2422	-245.1		PPB	21 Mar 2019 16:59:53	U	SMPL
K1934-02 - 1	EFF-WW	246	0.0060	-245.1		PPB	21 Mar 2019 17:02:09	U	SMPL
K1934-02DUP - 1	EFF-WWDUP	262	0.0099	-245.1		PPB	21 Mar 2019 17:04:26	U	SMPL
K1934-02MS - 1	EFF-WWMS	16475	3.9956	-245.1		PPB	21 Mar 2019 17:06:43	U	SMPL
K1934-02MSD - 1	EFF-WWMSD	16636	4.0352	-245.1		PPB	21 Mar 2019 17:09:00	U	SMPL
K1964-01 - 1	TW-11	30938	7.5511	-245.1		PPB	21 Mar 2019 17:11:16	U	SMPL
CCV30 - 1	CCV30	20697	5.0335	-245.1		PPB	21 Mar 2019 17:13:31	U	SMPL
CCB30 - 1	CCB30	67	-0.0380	-245.1		PPB	21 Mar 2019 17:15:48	U	SMPL
K1964-02 - 1	TW-11	212	-0.0023	-245.1		PPB	21 Mar 2019 17:18:03	U	SMPL
K1964-03 - 1	TW-4	313	0.0225	-245.1		PPB	21 Mar 2019 17:20:19	U	SMPL
K1964-04 - 1	TW-4	210	-0.0028	-245.1		PPB	21 Mar 2019 17:22:35	U	SMPL
K1964-05 - 1	TW-4	334	0.0276	-245.1		PPB	21 Mar 2019 17:24:51	U	SMPL
K1972-01 - 1	TW-8-031519	8116	1.9407	-245.1		PPB	21 Mar 2019 17:27:07	U	SMPL
K1972-02 - 1	TW-9-031519	3586	0.8271	-245.1		PPB	21 Mar 2019 17:29:23	U	SMPL
K1972-03 - 1	TW-10-031519	97993	24.0354	-245.1		PPB	21 Mar 2019 17:31:40	U	SMPL
K1972-06 - 1	TW-8-031519	9	-0.0522	-245.1		PPB	21 Mar 2019 17:33:57	U	SMPL
K1972-07 - 1	TW-9-031519	232	0.0026	-245.1		PPB	21 Mar 2019 17:36:16	U	SMPL
K1972-08 - 1	TW-10-031519	216	-0.0014	-245.1		PPB	21 Mar 2019 17:38:32	U	SMPL
CCV31 - 1	CCV31	20358	4.9502	-245.1		PPB	21 Mar 2019 17:40:48	U	SMPL
CCB31 - 1	CCB31	79	-0.0350	-245.1		PPB	21 Mar 2019 17:43:04	U	SMPL
K1934-02LX5 - 1		26	-0.0481	-245.1		PPB	21 Mar 2019 17:45:20	U	SMPL
K1934-02A - 1		16397	3.9765	-245.1		PPB	21 Mar 2019 17:47:36	U	SMPL
K1972-03DLX5 - 1		21068	5.1247	-245.1		PPB	21 Mar 2019 17:55:12	U	SMPL
CCV32 - 1	CCV32	20448	4.9723	-245.1		PPB	21 Mar 2019 17:57:27	U	SMPL
CCB32 - 1	CCB32	145	-0.0188	-245.1		PPB	21 Mar 2019 17:59:43	U	SMPL