



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/21/2019

Run Number: LB101542

BalanceID: WC SC-3

OvenID: WC Oven-2

FilterID: 15028095

ThermometerID: WET-OVEN 2

TEMP1 IN: 103 03/21/2019 14:00 TEMP OUT: 104 03/21/2019 15:20
TEMP2 IN: 104 03/21/2019 16:00 TEMP OUT: 104 03/22/2019 07:00
TEMP3 IN: 105 03/22/2019 11:30 TEMP OUT: 105 03/22/2019 13:10
TEMP4 IN: 105 03/22/2019 14:00 TEMP OUT: 105 03/22/2019 15:40

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB101542BL	LB101542BL	1.4253	1.4253	100	1.4253	1.4253	1.4253	0.0000	0
2	LB101542BS	LB101542BS	1.3962	1.3962	100	1.4463	1.4463	1.4463	0.0501	501
3	K2032-01	031819-TIPC-F-D-B	1.4339	1.4339	250	1.4482	1.4482	1.4482	0.0143	57.2
4	K2032-01DUP	031819-TIPC-F-D-BDUP	1.3752	1.3752	250	1.3895	1.3895	1.3895	0.0143	57.2
5	K2032-02	031819-TIPC-F-D-S	1.4075	1.4075	300	1.4231	1.4231	1.4231	0.0156	52
6	K2032-03	031819-TIPC-F-U-B	1.3973	1.3973	250	1.4089	1.4089	1.4089	0.0116	46.4
7	K2032-04	031819-TIPC-F-U-S	1.3982	1.3982	300	1.4101	1.4101	1.4101	0.0119	39.7
8	K2032-05	031819-LIFT-F-D-B	1.3988	1.3988	250	1.4064	1.4064	1.4064	0.0076	30.4
9	K2032-06	031819-LIFT-F-D-M	1.3748	1.3748	250	1.3894	1.3894	1.3894	0.0146	58.4
10	K2032-07	031819-LIFT-F-D-S	1.3886	1.3886	300	1.4025	1.4025	1.4025	0.0139	46.3
11	K2032-08	031819-LIFT-F-U-B	1.4230	1.4230	300	1.4444	1.4444	1.4444	0.0214	71.3
12	K2032-09	031819-LIFT-F-U-M	1.3745	1.3745	300	1.3876	1.3876	1.3876	0.0131	43.7
13	K2032-10	031819-LIFT-F-U-S	1.3953	1.3953	400	1.4088	1.4088	1.4088	0.0135	33.8
14	K2032-11	031819-DBRI-E-D-B	1.4060	1.4060	300	1.4235	1.4235	1.4235	0.0175	58.3
15	K2032-12	031819-DBRI-E-D-M	1.4231	1.4231	300	1.4373	1.4373	1.4373	0.0142	47.3



TOTAL SUSPENDED SOLIDS - SM2540D

TEMP1 IN: 103 03/21/2019 14:00 TEMP OUT: 104 03/21/2019 15:20
TEMP2 IN: 104 03/21/2019 16:00 TEMP OUT: 104 03/22/2019 07:00
TEMP3 IN: 105 03/22/2019 11:30 TEMP OUT: 105 03/22/2019 13:10
TEMP4 IN: 105 03/22/2019 14:00 TEMP OUT: 105 03/22/2019 15:40

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/21/2019

Run Number: LB101542

BalanceID: WC SC-3

OvenID: WC Oven-2

FilterID: 15028095

ThermometerID: WET-OVEN 2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	K2032-13	031819-DBRI-E-D-S	1.3975	1.3975	300	1.4118	1.4118	1.4118	0.0143	47.7
17	K2032-14	031819-DBRI-E-U-B	1.3811	1.3811	300	1.3961	1.3961	1.3961	0.0150	50
18	K2032-15	031819-DBRI-E-U-M	1.3996	1.3996	300	1.4126	1.4126	1.4126	0.0130	43.3
19	K2032-16	031819-DBRI-E-U-S	1.4163	1.4163	300	1.4262	1.4262	1.4262	0.0099	33
20	K2032-17	031819-JETT-E-D-B	1.3802	1.3802	400	1.3961	1.3961	1.3961	0.0159	39.8
21	K2032-18	031819-JETT-E-D-S	1.3791	1.3791	400	1.3922	1.3922	1.3922	0.0131	32.8
22	K2032-19	031819-JETT-E-U-B	1.3896	1.3896	300	1.4022	1.4022	1.4022	0.0126	42
23	K2032-20	031819-JETT-E-U-S	1.3882	1.3882	250	1.3964	1.3964	1.3964	0.0082	32.8
24	K2032-20DUP	031819-JETT-E-U-SDUP	1.4072	1.4072	250	1.4157	1.4157	1.4157	0.0085	34

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS032219

WorkList ID : 123766

Date : 03-22-2019 08:01:02

23/01/1542

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/25/2019	Water	K2032-01	TSS	Cool 4 deg C	HDR103	F22	031819-TIPC-F-D-B	03/18/2019	SM2540 D
03/25/2019	Water	K2032-02	TSS	Cool 4 deg C	HDR103	F22	031819-TIPC-F-D-S	03/18/2019	SM2540 D
03/25/2019	Water	K2032-03	TSS	Cool 4 deg C	HDR103	F22	031819-TIPC-F-U-B	03/18/2019	SM2540 D
03/25/2019	Water	K2032-04	TSS	Cool 4 deg C	HDR103	F22	031819-TIPC-F-U-S	03/18/2019	SM2540 D
03/25/2019	Water	K2032-05	TSS	Cool 4 deg C	HDR103	F22	031819-LIFT-F-D-B	03/18/2019	SM2540 D
03/25/2019	Water	K2032-06	TSS	Cool 4 deg C	HDR103	F22	031819-LIFT-F-D-M	03/18/2019	SM2540 D
03/25/2019	Water	K2032-07	TSS	Cool 4 deg C	HDR103	F22	031819-LIFT-F-D-S	03/18/2019	SM2540 D
03/25/2019	Water	K2032-08	TSS	Cool 4 deg C	HDR103	F22	031819-LIFT-F-U-B	03/18/2019	SM2540 D
03/25/2019	Water	K2032-09	TSS	Cool 4 deg C	HDR103	F22	031819-LIFT-F-U-M	03/18/2019	SM2540 D
03/25/2019	Water	K2032-10	TSS	Cool 4 deg C	HDR103	F22	031819-LIFT-F-U-S	03/18/2019	SM2540 D
03/25/2019	Water	K2032-11	TSS	Cool 4 deg C	HDR103	F22	031819-DBRI-E-D-B	03/18/2019	SM2540 D
03/25/2019	Water	K2032-12	TSS	Cool 4 deg C	HDR103	F22	031819-DBRI-E-D-M	03/18/2019	SM2540 D
03/25/2019	Water	K2032-13	TSS	Cool 4 deg C	HDR103	F22	031819-DBRI-E-D-S	03/18/2019	SM2540 D
03/25/2019	Water	K2032-14	TSS	Cool 4 deg C	HDR103	F22	031819-DBRI-E-U-B	03/18/2019	SM2540 D
03/25/2019	Water	K2032-15	TSS	Cool 4 deg C	HDR103	F22	031819-DBRI-E-U-M	03/18/2019	SM2540 D
03/25/2019	Water	K2032-16	TSS	Cool 4 deg C	HDR103	F22	031819-DBRI-E-U-S	03/18/2019	SM2540 D
03/25/2019	Water	K2032-17	TSS	Cool 4 deg C	HDR103	F22	031819-JETT-E-D-B	03/18/2019	SM2540 D
03/25/2019	Water	K2032-18	TSS	Cool 4 deg C	HDR103	F22	031819-JETT-E-D-S	03/18/2019	SM2540 D
03/25/2019	Water	K2032-19	TSS	Cool 4 deg C	HDR103	F22	031819-JETT-E-U-B	03/18/2019	SM2540 D
03/25/2019	Water	K2032-20	TSS	Cool 4 deg C	HDR103	F22	031819-JETT-E-U-S	03/18/2019	SM2540 D

Date/Time

Received by:

Relinquished by:

03/22/19, 08:10

AS



Date/Time

Received by:

Relinquished by:

03/22/19, 12:30

AS