



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 03/21/2019Run Number: LB101546BalanceID: WC SC-3OvenID: WC Oven-2FilterID: 15028095ThermometerID: WET-OVEN 2

TEMP1 IN:	<u>103</u>	<u>03/21/2019 14:00</u>	TEMP OUT:	<u>104</u>	<u>03/21/2019 15:20</u>
TEMP2 IN:	<u>104</u>	<u>03/21/2019 16:00</u>	TEMP OUT:	<u>104</u>	<u>03/22/2019 07:00</u>
TEMP3 IN:	<u>105</u>	<u>03/22/2019 08:00</u>	TEMP OUT:	<u>105</u>	<u>03/22/2019 10:40</u>
TEMP4 IN:	<u>105</u>	<u>03/22/2019 11:20</u>	TEMP OUT:	<u>105</u>	<u>03/22/2019 13:00</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB101546BL	LB101546BL	1.4052	1.4052	100	1.4052	1.4052	1.4052	0.0000	0
2	LB101546BS	LB101546BS	1.3956	1.3956	100	1.4460	1.4460	1.4460	0.0504	504
3	K2053-01	SPDES-1	1.3804	1.3804	900	1.3810	1.3810	1.3810	0.0006	0.7

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss032219-3

WorkList ID : 123808

Date : 03-22-2019 07:10:52

22/10/546

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/22/2019	Water	K2053-01	TSS	Cool 4 deg C	SKAN06	F51	SPDES-1	03/21/2019	SM2540 D

Date/Time 03/22/19, 07:00
 Received by: AS
 Relinquished by: AS

Date/Time 03/22/19, 08:36
 Received by: AS
 Relinquished by: AS