

LB101556


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58989



A= 0.0000e+000

B= 1.7081e-004

C= -1.1139e-001

Rho= 0.9995997

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.104	-0.104	43	0.000	43				
0.2	0.200	0.196	-0.004	1798	0.0 %	1798				
2.5	2.500	2.561	0.061	15645	0.0 %	15645				
5.0	5.000	5.193	0.193	31057	0.0 %	31057				
7.5	7.500	7.390	-0.110	43915	0.0 %	43915				
10.0	10.000	9.964	-0.036	58989	0.0 %	58989				

LB101556 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Mar 2019 14:28:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	43	-	0.00007471B	PPB		22 Mar 2019 15:20:49	S	Std
0.2 - 1	50-2	1798	-	0.20007471B	PPB		22 Mar 2019 15:23:07	S	Std
2.5 - 1	52-5	15645	-	2.50007471B	PPB		22 Mar 2019 15:25:24	S	Std
5.0 - 1	55-0	31057	-	5.00007471B	PPB		22 Mar 2019 15:27:41	S	Std
7.5 - 1	57-5	43915	-	7.50007471B	PPB		22 Mar 2019 15:29:58	S	Std
10.0 - 1	510-0	58989	-	10.00007471B	PPB		22 Mar 2019 15:32:15	S	Std
ICV66 - 1	ICV66	25127	4.1805	-7471B	PPB		22 Mar 2019 15:37:12	U	SMPL
ICB66 - 1	ICB66	76	-0.0984	-7471B	PPB		22 Mar 2019 15:39:27	U	SMPL
CCV37 - 1	CCV37	30545	5.1059	-7471B	PPB		22 Mar 2019 15:41:43	U	SMPL
CCB37 - 1	CCB37	0	-0.1114	-7471B	PPB		22 Mar 2019 15:43:58	U	SMPL
CRA - 1	CRA	1776	0.1920	-7471B	PPB		22 Mar 2019 15:46:15	U	SMPL
HighStd - 1	HighStd	59501	10.0519	-7471B	PPB		22 Mar 2019 15:48:30	U	SMPL
ChkStd - 1	ChkStd	42019	7.0658	-7471B	PPB		22 Mar 2019 15:50:46	U	SMPL
PB118143BL - 1	PBS	114	-0.0919	-7471B	PPB		22 Mar 2019 15:53:04	U	SMPL
PB118143BS - 1	LCSS	21804	3.6129	-7471B	PPB		22 Mar 2019 15:55:21	U	SMPL
K2002-01 - 1	S-1(2-10)-COMP	1428	0.1325	-7471B	PPB		22 Mar 2019 15:57:38	U	SMPL
K2002-01DUP - 1	S-1(2-10)-COMP DUP	1466	0.1390	-7471B	PPB		22 Mar 2019 15:59:55	U	SMPL
K2002-01MS - 1	S-1(2-10)-COMPMS	23206	3.8524	-7471B	PPB		22 Mar 2019 16:02:12	U	SMPL
K2002-01MSD - 1	S-1(2-10)-COMPMSD	23468	3.8971	-7471B	PPB		22 Mar 2019 16:04:30	U	SMPL
K2002-03 - 1	S-2(5-30)-COMP	167	-0.0829	-7471B	PPB		22 Mar 2019 16:06:45	U	SMPL
CCV38 - 1	CCV38	30532	5.1037	-7471B	PPB		22 Mar 2019 16:09:01	U	SMPL
CCB38 - 1	CCB38	-12	-0.1134	-7471B	PPB		22 Mar 2019 16:11:16	U	SMPL
K2003-01 - 1	S4A(5-30)COMP	1767	0.1904	-7471B	PPB		22 Mar 2019 16:13:32	U	SMPL
K2004-01 - 1	SB-20(10-11)	22102	3.6638	-7471B	PPB		22 Mar 2019 16:15:47	U	SMPL
K2004-02 - 1	SB-20(14-15)	1725	0.1833	-7471B	PPB		22 Mar 2019 16:18:03	U	SMPL
K2004-03 - 1	SB-19(7.5-8.5)	4783	0.7056	-7471B	PPB		22 Mar 2019 16:20:19	U	SMPL
K2004-04 - 1	SB-19(13-14)	3092	0.4167	-7471B	PPB		22 Mar 2019 16:22:35	U	SMPL
K2004-05 - 1	SB-24(7-8)	225074	38.3330	-7471B	PPB		22 Mar 2019 16:24:52	U	SMPL
K2004-06 - 1	SB-24(12-13)	1218	0.0967	-7471B	PPB		22 Mar 2019 16:27:09	U	SMPL
K2004-07 - 1	SB-19B(7-8)	10479	1.6785	-7471B	PPB		22 Mar 2019 16:29:28	U	SMPL
K2004-08 - 1	SB-19B(13.5-14.5)	1502	0.1452	-7471B	PPB		22 Mar 2019 16:31:46	U	SMPL
K2004-09 - 1	SB-22(7-8)	15549	2.5445	-7471B	PPB		22 Mar 2019 16:34:02	U	SMPL
CCV39 - 1	CCV39	30182	5.0439	-7471B	PPB		22 Mar 2019 16:36:18	U	SMPL
CCB39 - 1	CCB39	0	-0.1114	-7471B	PPB		22 Mar 2019 16:38:35	U	SMPL
K2004-10 - 1	SB-22(12-13)	3154	0.4273	-7471B	PPB		22 Mar 2019 16:40:55	U	SMPL
K2004-11 - 1	SB-23(0.5-1.5)	8298	1.3060	-7471B	PPB		22 Mar 2019 16:43:11	U	SMPL
K2004-12 - 1	SB-23(4-5)	43699	7.3527	-7471B	PPB		22 Mar 2019 16:45:26	U	SMPL
K2010-01 - 1	TP-1	31510	5.2708	-7471B	PPB		22 Mar 2019 16:47:42	U	SMPL
K2011-01 - 1	20190142-COMPOSITE-A	25221	4.1966	-7471B	PPB		22 Mar 2019 16:49:59	U	SMPL
K2018-01 - 1	SP-1-A	1358	0.1206	-7471B	PPB		22 Mar 2019 16:52:17	U	SMPL
K2018-03 - 1	SP-1-B	1296	0.1100	-7471B	PPB		22 Mar 2019 16:54:34	U	SMPL
K2018-05 - 1	SP-1-C	1615	0.1645	-7471B	PPB		22 Mar 2019 16:56:51	U	SMPL
PB118175BL - 1	PBS	18	-0.1083	-7471B	PPB		22 Mar 2019 16:59:08	U	SMPL
PB118175BS - 1	LCSS	21432	3.5494	-7471B	PPB		22 Mar 2019 17:01:24	U	SMPL
CCV40 - 1	CCV40	30864	5.1604	-7471B	PPB		22 Mar 2019 17:03:40	U	SMPL
CCB40 - 1	CCB40	-27	-0.1160	-7471B	PPB		22 Mar 2019 17:06:00	U	SMPL
K1690-01 - 1	HD-02-032019	1019	0.0627	-7471B	PPB		22 Mar 2019 17:08:16	U	SMPL
K1694-01 - 1	EO-02-032019-A	4209	0.6075	-7471B	PPB		22 Mar 2019 17:10:33	U	SMPL
K1694-03 - 1	EO-02-032019-B	3498	0.4861	-7471B	PPB		22 Mar 2019 17:12:49	U	SMPL
K1694-05 - 1	EO-02-032019-C	8735	1.3806	-7471B	PPB		22 Mar 2019 17:15:05	U	SMPL
K2024-01 - 1	RCA-08	6814	1.0525	-7471B	PPB		22 Mar 2019 17:17:21	U	SMPL
K2024-02 - 1	RCA-09	8549	1.3488	-7471B	PPB		22 Mar 2019 17:19:40	U	SMPL
K2040-01 - 1	S4B(2-10)	883	0.0394	-7471B	PPB		22 Mar 2019 17:21:57	U	SMPL
K2040-03 - 1	S4B(12-20)	381	-0.0463	-7471B	PPB		22 Mar 2019 17:24:15	U	SMPL
K2040-05 - 1	S4B(22-28)	399	-0.0432	-7471B	PPB		22 Mar 2019 17:26:33	U	SMPL
K2040-05DUP - 1	S4B(22-28)DUP	405	-0.0422	-7471B	PPB		22 Mar 2019 17:28:49	U	SMPL
CCV41 - 1	CCV41	30594	5.1143	-7471B	PPB		22 Mar 2019 17:31:06	U	SMPL
CCB41 - 1	CCB41	57	-0.1017	-7471B	PPB		22 Mar 2019 17:33:22	U	SMPL
K2040-05MS - 1	S4B(22-28)MS	23793	3.9526	-7471B	PPB		22 Mar 2019 17:35:38	U	SMPL
K2040-05MSD - 1	S4B(22-28)MSD	23665	3.9308	-7471B	PPB		22 Mar 2019 17:37:54	U	SMPL
K2050-01 - 1	CS-5	8530	1.3456	-7471B	PPB		22 Mar 2019 17:40:10	U	SMPL
K2050-02 - 1	CS-4	7673	1.1992	-7471B	PPB		22 Mar 2019 17:42:26	U	SMPL
K2050-03 - 1	CS-3	9449	1.5026	-7471B	PPB		22 Mar 2019 17:44:43	U	SMPL
K2050-04 - 1	CS-2	5188	0.7748	-7471B	PPB		22 Mar 2019 17:47:01	U	SMPL
K2050-05 - 1	CS-1	8629	1.3625	-7471B	PPB		22 Mar 2019 17:49:17	U	SMPL
K1926-01 - 1	SUMP-MUCK	1482	0.1417	-7471B	PPB		22 Mar 2019 17:51:35	U	SMPL
K2002-01LX5 - 1		253	-0.0682	-7471B	PPB		22 Mar 2019 17:53:52	U	SMPL
K2002-01A - 1		22459	3.7248	-7471B	PPB		22 Mar 2019 17:56:07	U	SMPL
CCV42 - 1	CCV42	30330	5.0692	-7471B	PPB		22 Mar 2019 17:58:24	U	SMPL
CCB42 - 1	CCB42	20	-0.1080	-7471B	PPB		22 Mar 2019 18:00:42	U	SMPL
K2040-05LX5 - 1		37	-0.1051	-7471B	PPB		22 Mar 2019 18:02:58	U	SMPL
K2040-05A - 1		23606	3.9207	-7471B	PPB		22 Mar 2019 18:05:15	U	SMPL
K2004-05DLX10 - 1		23299	3.8683	-7471B	PPB		22 Mar 2019 18:07:31	U	SMPL
CCV43 - 1	CCV43	28186	4.7030	-7471B	PPB		22 Mar 2019 18:09:48	U	SMPL
CCB43 - 1	CCB43	70	-0.0994	-7471B	PPB		22 Mar 2019 18:12:05	U	SMPL