



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB101565

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP72881
Residual chlorine std, 0PPM	WP72875
Residual chlorine std, 0.1PPM	WP72878
Residual chlorine LOQ, 0.2PPM	WP72877
Residual chlorine std, 0.2PPM	WP72879
Residual chlorine std, and CCV , 0.4PPM	WP72883
Residual chlorine std, 1.6PPM	WP72882
Residual chlorine ICV-LCS, 0.4PPM	WP72884
Residual chlorine LOD, 0.05PPM	WP72876
Total Chlorine Powder Pillows	W2249

Intercept: 0.0000

Slope: 1.0000

Regression: 1.000000

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	03/22/2019	11:18
2	CAL2	0.1	1	0.000	0.100	0.100	03/22/2019	11:19
3	CAL3	0.2	1	0.000	0.200	0.200	03/25/2019	11:20
4	CAL4	0.4	1	0.000	0.400	0.400	03/22/2019	11:20
5	CAL5	0.8	1	0.000	0.800	0.800	03/22/2019	11:21
6	CAL6	1.6	1	0.000	1.600	1.600	03/22/2019	11:22



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB101565

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.390	0.390	0.390	03/22/2019	11:23
2	ICB				1	0.000	0.000	0.000	0.000	03/22/2019	11:24
3	CCV1			0.4	1	0.000	0.400	0.400	0.400	03/22/2019	11:25
4	CCB1				1	0.000	0.000	0.000	0.000	03/22/2019	11:26
5	LB101565BL	50	50		1	0.000	0.000	0.000	0.000	03/22/2019	11:27
6	LB101565BS	50	50	0.4	1	0.000	0.390	0.390	0.390	03/22/2019	11:28
7	K1560-09	50	50		1	0.000	0.060	0.060	0.060	03/22/2019	11:29
8	CCV2			0.4	1	0.000	0.400	0.400	0.400	03/22/2019	11:29
9	CCB2				1	0.000	0.000	0.000	0.000	03/22/2019	11:30