

Analytical Summary Report

Analysis Method: 9038

ANALYST: AMANDEEP

Parameter: Sulfate

SUPERVISOR REVIEW BY: apatel

Run Number: LB101566

BalanceID: WC SC-4

Reagent/Standard	Lot/Log #
sulfate ICV 20ppm	WP72953
Sulfate LCS std, 20ppm	WP72954
sulfate 0.0ppm	WP72943
sulfate 5ppm	WP72945
sulfate 10ppm	WP72944
sulfate 20ppm	WP72947
sulfate 30ppm	WP72949
sulfate 40ppm	WP72951
sulfate CCV 20ppm	WP72952
Sulfate std, 15ppm	WP72946
Sulfate std, 35ppm	WP72950
Sulfate std, 25ppm	WP72948
Sulfate LOD std 1.0PPM	WP72955

Intercept: -2.776

Slope: 3.3359

Regression: 0.995246

Seq	Lab ID	True Val (mg/l)	DF	MATRIX	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	Water	0.302	0.441	0.139	03/26/2019	09:40
2	CAL2	5	1	Water	0.312	14.200	13.888	03/26/2019	09:44
3	CAL3	10	1	Water	0.309	34.200	33.891	03/26/2019	09:47
4	CAL4	15	1	Water	0.319	44.200	43.881	03/26/2019	09:50
5	CAL5	20	1	Water	0.299	59.200	58.901	03/26/2019	09:53
6	CAL6	25	1	Water	0.324	83.700	83.376	03/26/2019	09:57
7	CAL7	30	1	Water	0.281	91.100	90.819	03/26/2019	10:01
8	CAL8	35	1	Water	0.311	113.000	112.689	03/26/2019	10:06
9	CAL9	40	1	Water	0.308	138.200	137.892	03/26/2019	10:11



Analytical Summary Report

Analysis Method: 9038
Parameter: Sulfate
Run Number: LB101566

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	MATRIX	Initial Reading	Final Reading	Difference	Result (mg/L)	AnalDate	Anal Time
1	ICV			20	1	Water	0.300	59.100	58.800	18.459	03/26/2019	10:15
2	ICB				1	Water	0.302	0.416	0.114	0.866	03/26/2019	10:19
3	CCV1			20	1	Water	0.299	59.600	59.301	18.609	03/26/2019	10:24
4	CCB1				1	Water	0.270	0.376	0.106	0.864	03/26/2019	10:29
5	LB101566BL	5.01	50		1	Solid	0.310	0.426	0.116	0.867	03/26/2019	10:34
6	LB101566BS	5.02	50	20	1	Solid	0.315	58.600	58.285	18.304	03/26/2019	10:39
7	K1560-03	5.00	50		1	Solid	0.299	2.880	2.581	1.606	03/26/2019	10:45
8	CCV2			20	1	Water	0.329	60.000	59.671	18.720	03/26/2019	10:49
9	CCB2				1	Water	0.265	0.379	0.114	0.866	03/26/2019	10:55