



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/25/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 16:10
In Date: 03/22/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 03/23/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101576

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2076-01	MH0J61	1	1.00	2.00	2.00	100.0
K2076-02	MH0Z82	2	1.00	2.00	2.00	100.0
K2076-03	MH0Z83	3	1.00	2.00	2.00	100.0
K2076-04	MH0Z84	4	1.00	2.00	2.00	100.0
K2076-05	MH0Z85	5	1.00	2.00	2.00	100.0
K2076-06	MH0Z86	6	1.00	2.00	2.00	100.0
K2076-07	MH0Z87	7	1.00	2.00	2.00	100.0
K2076-08	MH0Z88	8	1.00	2.00	2.00	100.0
K2076-09	MH0Z89	9	1.00	2.00	2.00	100.0
K2076-10	MH0Z90	10	1.00	2.00	2.00	100.0
K2076-11	MH0Z91	11	1.00	2.00	2.00	100.0
K2076-12	MH0Z92	12	1.00	2.00	2.00	100.0
K2076-13	MH0Z93	13	1.00	2.00	2.00	100.0
K2076-14	MH0Z94	14	1.00	2.00	2.00	100.0
K2076-15	MH0Z95	15	1.00	2.00	2.00	100.0
K2076-16	MH0Z96	16	1.00	2.00	2.00	100.0
K2076-17	MH0Z97	17	1.00	2.00	2.00	100.0
K2076-18	MH0Z98	18	1.00	2.00	2.00	100.0
K2076-19	MH0Z98D	19	1.00	2.00	2.00	100.0
K2076-20	MH0Z98S	20	1.00	2.00	2.00	100.0
K2076-21	MH0Z99	21	1.00	2.00	2.00	100.0
K2076-22	MH0ZA0	22	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$