



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101578

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/22/2019	14:30
2	CAL2	1	Water	NA	NA	20.1	7.00	03/22/2019	14:31
3	CAL3	1	Water	NA	NA	20.1	10.02	03/22/2019	14:32
4	ICV	1	Water	NA	NA	20.1	7.01	03/22/2019	14:33
5	CCV1	1	Water	NA	NA	20.2	2.02	03/22/2019	14:35
6	K2072-01	1	Water	NA	NA	16.4	4.07	03/22/2019	14:36
7	K2072-01DUP	1	Water	NA	NA	16.5	4.08	03/22/2019	14:37
8	K2075-01	1	Water	NA	NA	11.9	11.60	03/22/2019	14:39
9	CCV2	1	Water	NA	NA	20.1	12.01	03/22/2019	14:40

LB 101578

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2075

WorkList ID : 123831

Date : 03-22-2019 14:16:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/18/2019	Water	K2072-01-1A	pH	Cool 4 deg C	STER15	F21	GILROY-OMEGA-CELL	03/18/2019	9040C
03/22/2019	Water	K2075-01-1A	pH	Cool 4 deg C	PSEG01	K31	NB-21752	03/22/2019	9040C

Date/Time 03-22-19 14:20

Received by: JDRelinquished by: CD

Date/Time

Received by:

Relinquished by:

03-22-19 15:20

JD

LD