



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101583

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/22/2019	17:00
2	CAL2	1	Water	NA	NA	20.2	7.01	03/22/2019	17:01
3	CAL3	1	Water	NA	NA	20.2	10.02	03/22/2019	17:02
4	ICV	1	Water	NA	NA	20.2	7.00	03/22/2019	17:05
5	CCV1	1	Water	NA	NA	20.2	2.01	03/22/2019	17:06
6	K2040-02	1	Solid	20.02	20	22.6	8.23	03/22/2019	17:07
7	K2040-02DUP	1	Solid	20.03	20	22.7	8.24	03/22/2019	17:08
8	K2040-04	1	Solid	20.05	20	22.6	8.71	03/22/2019	17:10
9	K2040-06	1	Solid	20.04	20	22.3	8.62	03/22/2019	17:11
10	K2058-02	1	Solid	20.02	20	21.8	7.60	03/22/2019	17:12
11	K2058-04	1	Solid	20.01	20	22.0	7.14	03/22/2019	17:13
12	K2061-02	1	Solid	20.03	20	21.6	10.48	03/22/2019	17:15
13	K2071-03	1	Solid	20.02	20	21.8	8.51	03/22/2019	17:16
14	K2071-04	1	Solid	20.03	20	22.0	9.23	03/22/2019	17:18
15	K2071-06	1	Solid	20.02	20	21.8	8.49	03/22/2019	17:20
16	CCV2	1	Water	NA	NA	20.1	12.01	03/22/2019	17:21

YB 101583

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2040

WorkList ID : 123828

Date : 03-22-2019 14:11:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2019	Solid	K2040-02	Corrosivity	Cool 4 deg C	SKAN04	F51	S4B(2-10)	03/20/2019	9045D
03/20/2019	Solid	K2040-04	Corrosivity	Cool 4 deg C	SKAN04	F51	S4B(12-20)	03/20/2019	9045D
03/20/2019	Solid	K2040-06	Corrosivity	Cool 4 deg C	SKAN04	F51	S4B(22-28)	03/20/2019	9045D
03/21/2019	Solid	K2058-02	Corrosivity	Cool 4 deg C	SKAN04	F61	S3B(5-15)COMP	03/21/2019	9045D
03/21/2019	Solid	K2058-04	Corrosivity	Cool 4 deg C	SKAN04	F61	S3B(17-25)COMP	03/21/2019	9045D
03/22/2019	Solid	K2061-02	Corrosivity	Cool 4 deg C	PSEG01	K31	26524-27245	03/22/2019	9045D
03/22/2019	Solid	K2071-03	Corrosivity	Cool 4 deg C	PSEG01		DRUM-1	03/22/2019	9045D
03/22/2019	Solid	K2071-04	Corrosivity	Cool 4 deg C	PSEG01		DRUM-2	03/22/2019	9045D
03/22/2019	Solid	K2071-06	Corrosivity	Cool 4 deg C	PSEG01	K31	DRUM-3	03/22/2019	9045D

Date/Time 03-22-19 16:00
 Received by: Jo
 Relinquished by: CO

Date/Time 03-22-19 17:30
 Received by: [Signature]
 Relinquished by: Jo