

TOTAL VOLATILE SOLIDS - 160.4

TEMP1 IN: 103 03/25/2019 15:00 **TEMP OUT:** 105 03/26/2019 07:00
TEMP2 IN: 105 03/26/2019 07:40 **TEMP OUT:** 105 03/26/2019 08:40
TEMP3 IN: 550 03/26/2019 09:20 **TEMP OUT:** 550 03/26/2019 11:05
TEMP4 IN: 550 03/26/2019 12:20 **TEMP OUT:** 550 03/26/2019 13:30

Run Number: LB101588

SUPERVISOR: apatel

ANALYST: Aparna

BalanceID: WC SC-3

OvenID: W-OVEN-2

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Empty Dish + Sample Weight (g)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result %
1	LB101588BL	56.9856	56.9856	56.9856	56.9856	56.9856	56.9856	56.9856	0	0
2	K2003-01	55.899	55.899	79.6766	77.6271	77.6271	77.527	77.527	0.1001	0.5
3	K2003-01DUP	57.1857	57.1857	81.119	78.983	78.983	78.8534	78.8534	0.1296	0.6
4	K2040-01	60.9426	60.9426	83.32	82.2892	82.2892	82.1635	82.1635	0.1257	0.6
5	K2040-03	59.3969	59.3969	82.043	79.5034	79.5034	79.399	79.399	0.1044	0.5
6	K2040-05	59.5646	59.5646	83.6709	80.6144	80.6144	80.512	80.512	0.1024	0.5
7	K2058-01	53.5539	53.5539	76.366	73.9869	73.9869	73.909	73.909	0.0779	0.4
8	K2058-03	60.8894	60.8894	84.217	81.3127	81.3127	81.2457	81.2457	0.067	0.3

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result mg/L = $\frac{D}{F - E} \star 100$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS032519

WorkList ID : 123872

Date : 03-25-2019 08:10:54

2B/10/588

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2019	Solid	K2003-01	TVS	Cool 4 deg C	SKAN04	F41	S4A(5-30)COMP	03/19/2019	160.4
03/27/2019	Solid	K2040-01	TVS	Cool 4 deg C	SKAN04	F51	S4B(2-10)	03/20/2019	160.4
03/27/2019	Solid	K2040-03	TVS	Cool 4 deg C	SKAN04	F51	S4B(12-20)	03/20/2019	160.4
03/27/2019	Solid	K2040-05	TVS	Cool 4 deg C	SKAN04	F51	S4B(22-28)	03/20/2019	160.4
03/28/2019	Solid	K2058-01	TVS	Cool 4 deg C	SKAN04	F61	S3B(5-15)COMP	03/21/2019	160.4
03/28/2019	Solid	K2058-03	TVS	Cool 4 deg C	SKAN04	F61	S3B(17-25)COMP	03/21/2019	160.4

Date/Time

Received by:

Relinquished by:

03/25/19, 14:00



Date/Time

Received by:

Relinquished by:

03/25/19, 16:00

