

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB101595

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP72857
calibration std. phosphate 0.5 ppm	WP72856
calibration std. phosphate 0.3 ppm	WP72855
calibration std. 0 ppm	WP72852
phosphate ICV std.	WP72861
phosphate CCV std.	WP72860
calibration std. phosphate 0.1 ppm	WP72854
calibration std. phosphate 0.05 ppm	WP72853
Combined reagent	WP72859
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: -0.0007Slope: 0.6498Regression: 0.999842

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	03/25/2019	13:08
2	CAL2	0.05	1	50	50	0.033	03/25/2019	13:08
3	CAL3	0.10	1	50	50	0.066	03/25/2019	13:09
4	CAL4	0.30	1	50	50	0.186	03/25/2019	13:09
5	CAL5	0.50	1	50	50	0.329	03/25/2019	13:10
6	CAL6	1.00	1	50	50	0.649	03/25/2019	13:11



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB101595

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.494	03/25/2019	13:11
2	ICB		1	50	50	0.000	0.001	03/25/2019	13:12
3	CCV1	0.50	1	50	50	0.328	0.506	03/25/2019	13:12
4	CCB1		1	50	50	0.000	0.001	03/25/2019	13:13
5	PB118247BL		1	50	50	0.000	0.001	03/25/2019	13:13
6	PB118247BS	0.50	1	50	50	0.296	0.457	03/25/2019	13:14
7	K1833-13		1	50	50	1.379	2.123	03/25/2019	13:15
8	K2041-01		1	50	50	0.014	0.023	03/25/2019	13:15
9	K2041-02		1	50	50	0.018	0.029	03/25/2019	13:16
10	K2041-02DUP		1	50	50	0.019	0.030	03/25/2019	13:16
11	K2041-02MS	0.50	1	50	50	0.330	0.509	03/25/2019	13:17
12	K2041-02MSD	0.50	1	50	50	0.331	0.510	03/25/2019	13:18
13	K2067-01		1	50	50	3.230	4.972	03/25/2019	13:18
14	K1833-13		10	50	50	0.443	0.683	03/25/2019	13:40
15	CCV2	0.50	1	50	50	0.329	0.507	03/25/2019	13:41
16	CCB2		1	50	50	0.000	0.001	03/25/2019	13:42
17	K2067-01		50	50	50	0.376	0.580	03/25/2019	13:43
18	CCV3	0.50	1	50	50	0.327	0.504	03/25/2019	13:44
19	CCB3		1	50	50	0.000	0.001	03/25/2019	13:44