



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/26/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:10
In Date: 03/25/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 03/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101607

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1693-13	OR-01-032519-A	1	1.13	9.6	8.7	89.4
K1693-15	OR-01-032519-B	2	1.16	9.79	8.87	89.3
K1950-01	FK-SF-03-005-A	3	1.15	9.62	8.9	91.5
K1950-02	FK-SF-03-005-A	4	1.12	9.51	8.84	92.0
K1950-03	FK-SF-03-005-B	5	1.16	9.7	8.93	91.0
K1950-04	FK-SF-03-005-B	6	1.15	9.66	8.98	92.0
K2015-05	PST-072-B156-031919	7	1.18	9.98	9.36	93.0
K2081-01	S3A(2-10)	8	1.13	9.53	8.95	93.1
K2081-03	S3A(12-22)	9	1.12	9.54	8.27	84.9
K2081-05	S3A(10)	10	1.17	9.67	9.19	94.4
K2081-07	S3A(22)	11	1.18	9.71	8.54	86.3
K2081-14	S3A(22-28)-SB(27) COMP	12	1.16	9.81	8.93	89.8
K2081-16	S4C(6-18) COMP	13	1.14	9.62	8.86	91.0
K2081-18	S4C(18-29) COMP	14	1.15	9.69	8.91	90.9
K2081-20	S3B(27)	15	1.15	9.55	8.47	87.1
K2081-22	S4C(12)	16	1.11	9.6	8.83	90.9
K2081-24	S4C(21)	17	1.13	9.95	9.02	89.5
K2082-04	SVOC-GPC-BLANK	18	1.00	2.00	2.00	100.0
K2082-05	PEST-GPC-BLANK	19	1.00	2.00	2.00	100.0
K2082-06	PEST-GPC-BLANK-SPIKE	20	1.00	2.00	2.00	100.0
K2082-07	PCB-GPC-BLANK	21	1.00	2.00	2.00	100.0
K2082-08	PCB-GPC-BLANK-SPIKE	22	1.00	2.00	2.00	100.0
K2082-09	SVOC-GPC2-BLANK	23	1.00	2.00	2.00	100.0
K2082-10	PEST-GPC2-BLANK	24	1.00	2.00	2.00	100.0
K2082-11	PEST-GPC2-BLANK-SPIKE	25	1.00	2.00	2.00	100.0
K2082-12	PCB-GPC2-BLANK	26	1.00	2.00	2.00	100.0
K2082-13	PCB-GPC2-BLANK-SPIKE	27	1.00	2.00	2.00	100.0
K2091-01	DRUM-COMP	28	1.14	9.97	8.53	83.7
K2095-01	72-11976	29	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$