

LB101615


245.1

INSTRUMENT ID: C01

Linear

μ Abs.:

4264E



A= 0.0000e+000

B= 2.3604e-004

C= -2.8460e-002

Rho= 0.9999063

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.037	-0.037	-35	0.000	-35				
0.05	0.050	0.041	-0.009	294	0.0 %	294				
0.2	0.200	0.219	0.019	1047	0.0 %	1047				
2.5	2.500	2.536	0.036	10863	0.0 %	10863				
5.0	5.000	5.056	0.056	21542	0.0 %	21542				
7.5	7.500	7.398	-0.102	31462	0.0 %	31462				
10.0	10.000	10.038	0.038	42646	0.0 %	42646				

LB101615 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 26 Mar 2019 08:38:24				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	-35	-	0.0000245.1	PPB	26 Mar 2019 10:13:30	S	Std
0.05 - 1	50-05	294	-	0.0500245.1	PPB	26 Mar 2019 10:15:48	S	Std
0.2 - 1	50-2	1047	-	0.2000245.1	PPB	26 Mar 2019 10:18:05	S	Std
2.5 - 1	52-5	10863	-	2.5000245.1	PPB	26 Mar 2019 10:20:22	S	Std
5.0 - 1	55-0	21542	-	5.0000245.1	PPB	26 Mar 2019 10:22:39	S	Std
7.5 - 1	57-5	31462	-	7.5000245.1	PPB	26 Mar 2019 10:24:55	S	Std
10.0 - 1	510-0	42646	-	10.0000245.1	PPB	26 Mar 2019 10:47:05	S	Std
ICV68 - 1	ICV68	16003	3.7488	-245.1	PPB	26 Mar 2019 10:50:02	U	SMPL
ICB68 - 1	ICB68	-46	-0.0393	-245.1	PPB	26 Mar 2019 10:52:17	U	SMPL
CCV46 - 1	CCV46	21884	5.1370	-245.1	PPB	26 Mar 2019 10:54:32	U	SMPL
CCB46 - 1	CCB46	159	0.0091	-245.1	PPB	26 Mar 2019 10:56:47	U	SMPL
CRA - 1	CRA	281	0.0379	-245.1	PPB	26 Mar 2019 10:59:02	U	SMPL
HighStd - 1	HighStd	42978	10.1159	-245.1	PPB	26 Mar 2019 11:01:18	U	SMPL
ChkStd - 1	ChkStd	29543	6.9448	-245.1	PPB	26 Mar 2019 11:03:33	U	SMPL
PB118258BL - 1	PBW	-25	-0.0344	-245.1	PPB	26 Mar 2019 11:05:50	U	SMPL
PB118258BS - 1	LCSW	17926	4.2027	-245.1	PPB	26 Mar 2019 11:08:06	U	SMPL
K1981-02 - 1	METALS-2019	592	0.1113	-245.1	PPB	26 Mar 2019 11:10:23	U	SMPL
K1981-02DUP - 1	METALS-2019DUP	585	0.1096	-245.1	PPB	26 Mar 2019 11:12:40	U	SMPL
K1981-02MS - 1	METALS-2019MS	14885	3.4849	-245.1	PPB	26 Mar 2019 11:14:57	U	SMPL
K1981-02MSD - 1	METALS-2019MSD	14570	3.4106	-245.1	PPB	26 Mar 2019 11:17:14	U	SMPL
K2033-01 - 1	032019-LIFT-F-D-B	-77	-0.0466	-245.1	PPB	26 Mar 2019 11:19:31	U	SMPL
CCV47 - 1	CCV47	21205	4.9767	-245.1	PPB	26 Mar 2019 11:21:48	U	SMPL
CCB47 - 1	CCB47	109	-0.0027	-245.1	PPB	26 Mar 2019 11:24:03	U	SMPL
K2033-02 - 1	032019-LIFT-F-D-M	30	-0.0214	-245.1	PPB	26 Mar 2019 11:26:19	U	SMPL
K2033-03 - 1	032019-LIFT-F-D-S	39	-0.0193	-245.1	PPB	26 Mar 2019 11:28:34	U	SMPL
K2033-04 - 1	032019-LIFT-F-U-B	19	-0.0240	-245.1	PPB	26 Mar 2019 11:30:50	U	SMPL
K2033-05 - 1	032019-LIFT-F-U-M	29	-0.0216	-245.1	PPB	26 Mar 2019 11:33:06	U	SMPL
K2033-06 - 1	032019-LIFT-F-U-S	23	-0.0230	-245.1	PPB	26 Mar 2019 11:35:22	U	SMPL
K2034-01 - 1	031619-LIFT-E-D-B	28	-0.0219	-245.1	PPB	26 Mar 2019 11:37:38	U	SMPL
K2034-02 - 1	031619-LIFT-E-D-M	47	-0.0174	-245.1	PPB	26 Mar 2019 11:39:55	U	SMPL
K2034-03 - 1	031619-LIFT-E-D-S	32	-0.0209	-245.1	PPB	26 Mar 2019 11:42:12	U	SMPL
K2034-04 - 1	031619-LIFT-E-U-B	35	-0.0202	-245.1	PPB	26 Mar 2019 11:44:29	U	SMPL
K2034-05 - 1	031619-LIFT-E-U-M	34	-0.0204	-245.1	PPB	26 Mar 2019 11:46:46	U	SMPL
CCV48 - 1	CCV48	21834	5.1252	-245.1	PPB	26 Mar 2019 11:49:02	U	SMPL
CCB48 - 1	CCB48	157	0.0086	-245.1	PPB	26 Mar 2019 11:51:18	U	SMPL
K2034-06 - 1	031619-LIFT-E-U-S	23	-0.0230	-245.1	PPB	26 Mar 2019 11:53:34	U	SMPL
K2035-11 - 1	031919-LIFT-F-D-B	72	-0.0115	-245.1	PPB	26 Mar 2019 11:55:49	U	SMPL
K2035-12 - 1	031919-LIFT-F-D-M	61	-0.0141	-245.1	PPB	26 Mar 2019 11:58:05	U	SMPL
K2035-13 - 1	031919-LIFT-F-D-S	62	-0.0138	-245.1	PPB	26 Mar 2019 12:00:21	U	SMPL
K2035-14 - 1	031919-LIFT-F-U-B	33	-0.0207	-245.1	PPB	26 Mar 2019 12:02:37	U	SMPL
K2035-15 - 1	031919-LIFT-F-U-M	34	-0.0204	-245.1	PPB	26 Mar 2019 12:04:54	U	SMPL
K2035-16 - 1	031919-LIFT-F-U-S	46	-0.0176	-245.1	PPB	26 Mar 2019 12:07:10	U	SMPL
K2038-01 - 1	37-ST-2019	25	-0.0226	-245.1	PPB	26 Mar 2019 12:09:28	U	SMPL
PB118259BL - 1	PBW	-3	-0.0292	-245.1	PPB	26 Mar 2019 12:11:46	U	SMPL
PB118259BS - 1	LCSW	17779	4.1680	-245.1	PPB	26 Mar 2019 12:14:02	U	SMPL
CCV49 - 1	CCV49	21364	5.0142	-245.1	PPB	26 Mar 2019 12:16:18	U	SMPL
CCB49 - 1	CCB49	148	0.0065	-245.1	PPB	26 Mar 2019 12:18:34	U	SMPL
K1983-01 - 1	EFFLUENT-DAY1-METALS	135	0.0034	-245.1	PPB	26 Mar 2019 12:20:51	U	SMPL
K1983-01DUP - 1	EFFLUENT-DAY1-METALS DUP	66	-0.0129	-245.1	PPB	26 Mar 2019 12:23:07	U	SMPL
K1983-01MS - 1	EFFLUENT-DAY1-METALS MS	16853	3.9495	-245.1	PPB	26 Mar 2019 12:25:23	U	SMPL
K1983-01MSD - 1	EFFLUENT-DAY1-METALS MSD	16721	3.9183	-245.1	PPB	26 Mar 2019 12:27:39	U	SMPL
K2032-05 - 1	031819-LIFT-F-D-B	-43	-0.0386	-245.1	PPB	26 Mar 2019 12:29:56	U	SMPL
K2032-06 - 1	031819-LIFT-F-D-M	41	-0.0188	-245.1	PPB	26 Mar 2019 12:32:13	U	SMPL
K2032-07 - 1	031819-LIFT-F-D-S	50	-0.0167	-245.1	PPB	26 Mar 2019 12:34:30	U	SMPL
K2032-08 - 1	031819-LIFT-F-U-B	62	-0.0138	-245.1	PPB	26 Mar 2019 12:36:47	U	SMPL
K2032-09 - 1	031819-LIFT-F-U-M	28	-0.0219	-245.1	PPB	26 Mar 2019 12:39:05	U	SMPL
K2032-10 - 1	031819-LIFT-F-U-S	27	-0.0221	-245.1	PPB	26 Mar 2019 12:41:21	U	SMPL
CCV50 - 1	CCV50	21122	4.9571	-245.1	PPB	26 Mar 2019 12:43:36	U	SMPL
CCB50 - 1	CCB50	116	-0.0011	-245.1	PPB	26 Mar 2019 12:45:52	U	SMPL
K1981-02LX5 - 1		161	0.0095	-245.1	PPB	26 Mar 2019 12:48:08	U	SMPL
K1981-02A - 1		14856	3.4781	-245.1	PPB	26 Mar 2019 12:50:23	U	SMPL
K1983-01LX5 - 1		-109	-0.0542	-245.1	PPB	26 Mar 2019 12:52:40	U	SMPL
K1983-01A - 1		16634	3.8978	-245.1	PPB	26 Mar 2019 12:54:56	U	SMPL
CCV51 - 1	CCV51	21039	4.9375	-245.1	PPB	26 Mar 2019 12:57:13	U	SMPL
CCB51 - 1	CCB51	80	-0.0096	-245.1	PPB	26 Mar 2019 12:59:29	U	SMPL