



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB101627

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP72997
Turbidity std, 40NTU	WP72991
turbidity std 5 NTU	WP72994
turbidity std 20 NTU	WP72992
Turbidity std 1 NTU	WP72995
turbidity std 0 NTU	WP72996
turbidity std 10 NTU	WP72993

Intercept: 0.0565

Slope: 0.9683

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.269	03/26/2019	12:33
2	CAL2	1	1	1.030	03/26/2019	12:03
3	CAL3	5	1	4.690	03/26/2019	12:05
4	CAL4	10	1	9.840	03/26/2019	12:10
5	CAL5	20	1	19.200	03/26/2019	12:14
6	CAL6	40	1	38.900	03/26/2019	12:17



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB101627

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.890	10.155	03/26/2019	12:22
2	ICB		1	0.278	0.229	03/26/2019	12:26
3	CCV1	10	1	9.860	10.124	03/26/2019	12:34
4	CCB1		1	0.285	0.236	03/26/2019	12:39
5	LB101627BL		1	0.280	0.231	03/26/2019	12:45
6	K1560-07		1	0.869	0.839	03/26/2019	12:49
7	K1560-08		1	2.420	2.441	03/26/2019	12:55
8	K1833-38		1	9.590	9.846	03/26/2019	12:59
9	K1833-38DUP		1	9.670	9.928	03/26/2019	13:06
10	CCV2	10	1	9.920	10.186	03/26/2019	13:13
11	CCB2		1	0.285	0.236	03/26/2019	13:17