



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB101628

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP72997
Turbidity std, 40NTU	WP72991
turbidity std 5 NTU	WP72994
turbidity std 20 NTU	WP72992
Turbidity std 1 NTU	WP72995
turbidity std 0 NTU	WP72996
turbidity std 10 NTU	WP72993
TURBIDITY LOD STD, 0.5NTU	WP72998
TURBIDITY LOQ STD, 2.0NTU	WP72999

Intercept: 0.0565 Slope: 0.9683 Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.269	03/26/2019	12:00
2	CAL2	1	1	1.030	03/26/2019	12:03
3	CAL3	5	1	4.690	03/26/2019	12:05
4	CAL4	10	1	9.840	03/26/2019	12:10
5	CAL5	20	1	19.200	03/26/2019	12:14
6	CAL6	40	1	38.900	03/26/2019	12:17



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB101628

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.890	10.155	03/26/2019	12:22
2	ICB		1	0.278	0.229	03/26/2019	12:26
3	CCV1	10	1	9.920	10.186	03/26/2019	13:13
4	CCB1		1	0.285	0.236	03/26/2019	13:17
5	LB101628BL		1	0.272	0.223	03/26/2019	13:18
6	K1833-37		1	9.600	9.856	03/26/2019	13:20
7	K1833-37DUP		1	9.650	9.908	03/26/2019	13:25
8	CCV2	10	1	9.460	9.711	03/26/2019	13:29
9	CCB2		1	0.265	0.215	03/26/2019	13:35