



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/27/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 16:30
In Date: 03/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:53
Out Date: 03/27/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101635

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2068-01	BF5M0	1	1.13	9.57	8.56	88.0
K2068-02	BF5M1	2	1.12	9.73	7.6	75.3
K2068-03	BF5M2	3	1.12	9.68	7.34	72.7
K2068-04	BF5M3	4	1.12	9.97	7.76	75.0
K2068-05	BF5M7	5	1.13	9.96	7.89	76.6
K2068-06	BF5M8	6	1.12	9.78	7.61	74.9
K2068-07	BFC97	7	1.13	9.83	7.42	72.3
K2068-08	BFC98	8	1.12	9.98	7.8	75.4
K2068-09	VHBLK01	9	1.00	2.00	2.00	100.0
K2068-10	BF5K6	10	1.13	9.89	8.89	88.6
K2068-11	BF5K7	11	1.14	9.77	7.74	76.5
K2068-12	BF5K8	12	1.14	9.97	7.56	72.7
K2068-13	BF5K9	13	1.13	9.63	8.18	82.9
K2068-14	BF5K9MS	14	1.13	9.63	8.18	82.9
K2068-15	BF5K9MSD	15	1.13	9.63	8.18	82.9
K2068-16	BF5L0	16	1.12	9.62	7.00	69.2
K2068-17	BF5L1	17	1.13	9.59	6.66	65.4
K2068-18	BF5M4	18	1.13	9.73	8.16	81.7
K2068-19	BF5J7	19	1.11	9.78	7.96	79.0
K2068-20	BF5J8	20	1.12	9.68	7.59	75.6
K2068-21	BF5J9	21	1.12	9.79	5.91	55.2
K2068-22	BF5K3	22	1.11	9.79	8.02	79.6
K2068-23	BF5K4	23	1.11	9.98	7.91	76.7
K2068-24	BF5K5	24	1.13	9.71	7.24	71.2

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$