

LB101636


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

59389



A= 0.0000e+000

B= 1.6997e-004

C= -5.4343e-002

Rho= 0.9999737

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	409	0.000	409				
0.2	0.200	0.221	0.021	1621	0.0 %	1621				
2.5	2.500	2.483	-0.017	14927	0.0 %	14927				
5.0	5.000	4.969	-0.031	29553	0.0 %	29553				
7.5	7.500	7.472	-0.028	44284	0.0 %	44284				
10.0	10.000	10.040	0.040	59389	0.0 %	59389				

LB101636 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Mar 2019 16:25:44

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	SD-D	409	-	0.00007471B	PPB	26 Mar 2019 16:27:10	S	Std
0.2 - 1	SD-2	1621	-	0.20007471B	PPB	26 Mar 2019 16:29:27	S	Std
2.5 - 1	S2-S	14927	-	2.50007471B	PPB	26 Mar 2019 16:31:45	S	Std
5.0 - 1	SC-0	29553	-	5.00007471B	PPB	26 Mar 2019 16:34:02	S	Std
7.5 - 1	S7-S	44284	-	7.50007471B	PPB	26 Mar 2019 16:36:18	S	Std
10.0 - 1	SD-0	59389	-	10.00007471B	PPB	26 Mar 2019 16:38:35	S	Std
ICV70 - 1	ICV70	23896	4.0072	-7471B	PPB	26 Mar 2019 16:42:08	U	SMPL
ICB70 - 1	ICB70	251	-0.0117	-7471B	PPB	26 Mar 2019 16:44:23	U	SMPL
CCV56 - 1	CCV56	29734	4.9994	-7471B	PPB	26 Mar 2019 16:46:39	U	SMPL
CCB56 - 1	CCB56	28	-0.0496	-7471B	PPB	26 Mar 2019 16:48:54	U	SMPL
CRA - 1	CRA	1556	0.2101	-7471B	PPB	26 Mar 2019 16:51:10	U	SMPL
HighStd - 1	HighStd	59835	10.1156	-7471B	PPB	26 Mar 2019 16:53:25	U	SMPL
ChkStd - 1	ChkStd	41413	6.9845	-7471B	PPB	26 Mar 2019 16:55:41	U	SMPL
PB118271BL - 1	PBS	-122	-0.0751	-7471B	PPB	26 Mar 2019 16:57:59	U	SMPL
PB118271BS - 1	LCSS	21010	3.5167	-7471B	PPB	26 Mar 2019 17:00:16	U	SMPL
K1693-13 - 1	OR-01-032519-A	2405	0.3544	-7471B	PPB	26 Mar 2019 17:02:32	U	SMPL
K1693-15 - 1	OR-01-032519-B	1965	0.2796	-7471B	PPB	26 Mar 2019 17:04:49	U	SMPL
K1950-01 - 1	FK-SF-03-005-A	6320	1.0198	-7471B	PPB	26 Mar 2019 17:07:07	U	SMPL
K1950-03 - 1	FK-SF-03-005-B	8489	1.3885	-7471B	PPB	26 Mar 2019 17:09:23	U	SMPL
K2058-01 - 1	S3B(5-15)COMP	402	0.0140	-7471B	PPB	26 Mar 2019 17:11:39	U	SMPL
CCV57 - 1	CCV57	29982	5.0416	-7471B	PPB	26 Mar 2019 17:13:54	U	SMPL
CCB57 - 1	CCB57	41	-0.0474	-7471B	PPB	26 Mar 2019 17:16:09	U	SMPL
K2058-03 - 1	S3B(17-25)COMP	237	-0.0141	-7471B	PPB	26 Mar 2019 17:18:30	U	SMPL
K2061-01 - 1	26524-27245	658	0.0575	-7471B	PPB	26 Mar 2019 17:20:45	U	SMPL
K2071-01 - 1	DRUM-1	2210	0.3213	-7471B	PPB	26 Mar 2019 17:23:01	U	SMPL
K2078-01 - 1	FK-SF-01-005-A	5947	0.9564	-7471B	PPB	26 Mar 2019 17:25:17	U	SMPL
K2078-03 - 1	FK-SF-01-005-B	7298	1.1861	-7471B	PPB	26 Mar 2019 17:27:33	U	SMPL
K2081-01 - 1	S3A(2-10)	640	0.0544	-7471B	PPB	26 Mar 2019 17:29:49	U	SMPL
K2081-03 - 1	S3A(12-22)	233	-0.0147	-7471B	PPB	26 Mar 2019 17:32:05	U	SMPL
K2081-14 - 1	S3A(22-28)-SB(27)COMP	152	-0.0285	-7471B	PPB	26 Mar 2019 17:34:22	U	SMPL
K2081-16 - 1	S4C(6-18)COMP	449	0.0220	-7471B	PPB	26 Mar 2019 17:36:40	U	SMPL
K2081-18 - 1	S4C(18-29)COMP	195	-0.0212	-7471B	PPB	26 Mar 2019 17:38:55	U	SMPL
CCV58 - 1	CCV58	30331	5.1009	-7471B	PPB	26 Mar 2019 17:41:12	U	SMPL
CCB58 - 1	CCB58	77	-0.0413	-7471B	PPB	26 Mar 2019 17:43:27	U	SMPL
K2091-01 - 1	DRUM-COMP	7749	1.2627	-7471B	PPB	26 Mar 2019 17:45:44	U	SMPL
K2091-01DUP - 1	DRUM-COMPDUP	7702	1.2547	-7471B	PPB	26 Mar 2019 17:48:02	U	SMPL
K2091-01MS - 1	DRUM-COMPMS	29434	4.9484	-7471B	PPB	26 Mar 2019 17:50:18	U	SMPL
K2091-01MSD - 1	DRUM-COMPMSD	29064	4.8856	-7471B	PPB	26 Mar 2019 17:52:34	U	SMPL
K2091-01LX5 - 1		1730	0.2397	-7471B	PPB	26 Mar 2019 17:54:50	U	SMPL
K2091-01A - 1		28658	4.8166	-7471B	PPB	26 Mar 2019 17:57:10	U	SMPL
CCV59 - 1	CCV59	31699	5.3334	-7471B	PPB	26 Mar 2019 17:59:26	U	SMPL
CCB59 - 1	CCB59	-5	-0.0552	-7471B	PPB	26 Mar 2019 18:01:44	U	SMPL