



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101639

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	03/25/2019	10:25
2	CAL2	1	Water	NA	NA	21.1	7.01	03/25/2019	10:26
3	CAL3	1	Water	NA	NA	21.1	10.04	03/25/2019	10:27
4	ICV	1	Water	NA	NA	21.1	7.00	03/25/2019	10:29
5	CCV1	1	Water	NA	NA	21.1	2.01	03/25/2019	10:30
6	K2056-01	1	Solid	20.03	20	22.7	6.70	03/25/2019	10:31
7	K2056-01DUP	1	Solid	20.04	20	22.8	6.71	03/25/2019	10:32
8	K2056-02	1	Solid	20.02	20	22.5	7.29	03/25/2019	10:34
9	K2056-03	1	Solid	20.04	20	21.9	8.01	03/25/2019	10:35
10	K2056-04	1	Solid	20.01	20	22.2	6.76	03/25/2019	10:36
11	CCV2	1	Water	NA	NA	21.2	12.01	03/25/2019	10:37

WORKLIST(Hardcopy Internal Chain)

28 10/639

WorkList Name : PH K2056

WorkList ID : 124364

Date : 03-25-2019 10:10:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/21/2019	Solid	K2056-01	pH	Cool 4 deg C	TRAC01	F42	CS-7	03/21/2019	9045D
03/21/2019	Solid	K2056-02	pH	Cool 4 deg C	TRAC01	F42	CS-8	03/21/2019	9045D
03/21/2019	Solid	K2056-03	pH	Cool 4 deg C	TRAC01	F42	CS-9	03/21/2019	9045D
03/21/2019	Solid	K2056-04	pH	Cool 4 deg C	TRAC01	F42	CS-10	03/21/2019	9045D

Date/Time 03-25-19 10-20

Received by: J0

Relinquished by: C0

Date/Time 03-25-19 11-25

Received by: J0

Relinquished by: C0