

LB101640


ISM02.4

INSTRUMENT ID : CUI

Linear

 μ Abs.:

41801



A= 0.0000e+000

B= 2.3245e-004

C= -5.6481e-002

Rho= 0.9983336

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% Δ
0.00	0.000	-0.040	-0.040	73	0.000	73					—
0.2	0.200	0.162	-0.038	939	0.0 %	939					-19
2.5	2.500	2.393	-0.107	10539	0.0 %	10539					-4
5.0	5.000	5.213	0.213	22667	0.0 %	22667					4
7.5	7.500	7.812	0.312	33848	0.0 %	33848					4
10.0	10.000	9.660	-0.340	41801	0.0 %	41801					-3

RS
03-26-19

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	73	-		0 ISM02.4	PPB	3/26/2019 18:17	Std
	0.2 S0.2	939	-		0.2 ISM02.4	PPB	3/26/2019 18:19	Std
	2.5 S2.5	10539	-		2.5 ISM02.4	PPB	3/26/2019 18:21	Std
	5 S5.0	22667	-		5 ISM02.4	PPB	3/26/2019 18:23	Std
	7.5 S7.5	33848	-		7.5 ISM02.4	PPB	3/26/2019 18:26	Std
	10 S10.0	41801	-		10 ISM02.4	PPB	3/26/2019 18:36	Std
ICV	ICV	18096	4.15	-	ISM02.4	PPB	3/26/2019 18:41	SMPL
ICB	ICB	-5	-0.0576	-	ISM02.4	PPB	3/26/2019 18:43	SMPL
CCV060	CCV060	22826	5.2495	-	ISM02.4	PPB	3/26/2019 18:45	SMPL
CCB060	CCB060	-9	-0.0586	-	ISM02.4	PPB	3/26/2019 18:47	SMPL
PB118262BL	PBW262	151	-0.0214	-	ISM02.4	PPB	3/26/2019 18:50	SMPL
K2021-21	MBF5Y0	108	-0.0314	-	ISM02.4	PPB	3/26/2019 18:52	SMPL
PB118263BL	PBW263	145	-0.0228	-	ISM02.4	PPB	3/26/2019 18:54	SMPL
K1889-01	MA41D4	148	-0.0221	-	ISM02.4	PPB	3/26/2019 18:56	SMPL
K1889-02	MA41E1	154	-0.0207	-	ISM02.4	PPB	3/26/2019 18:59	SMPL
K1889-03	MA41E1D	138	-0.0244	-	ISM02.4	PPB	3/26/2019 19:01	SMPL
K1889-04	MA41E1S	4908	1.0844	-	ISM02.4	PPB	3/26/2019 19:03	SMPL
K1889-05	MA41F2	227	-0.0037	-	ISM02.4	PPB	3/26/2019 19:05	SMPL
K1889-06	MA41N0	223	-0.0046	-	ISM02.4	PPB	3/26/2019 19:08	SMPL
K1889-08	MA41A9	17213	3.9447	-	ISM02.4	PPB	3/26/2019 19:10	SMPL
K1889-08	MA41A9	17048	3.9064	-	ISM02.4	PPB	3/26/2019 19:12	SMPL
K1889-08	MA41A9	17228	3.9482	-	ISM02.4	PPB	3/26/2019 19:14	SMPL
K1889-09	MA41G1	143	-0.0232	-	ISM02.4	PPB	3/26/2019 19:17	SMPL
K1889-10	MA41G4	152	-0.0211	-	ISM02.4	PPB	3/26/2019 19:19	SMPL
K1889-11	MA41G6	232	-0.0026	-	ISM02.4	PPB	3/26/2019 19:21	SMPL
K1889-12	MA41H2	439	0.0456	-	ISM02.4	PPB	3/26/2019 19:24	SMPL
K1889-13	MA41N2	241	-0.0005	-	ISM02.4	PPB	3/26/2019 19:26	SMPL
K1889-14	MA41G2	207	-0.0084	-	ISM02.4	PPB	3/26/2019 19:28	SMPL
K1889-15	MA41L6	257	0.0033	-	ISM02.4	PPB	3/26/2019 19:30	SMPL
K1889-16	MA41L6D	283	0.0093	-	ISM02.4	PPB	3/26/2019 19:33	SMPL
K1889-17	MA41L6S	5622	1.2504	-	ISM02.4	PPB	3/26/2019 19:35	SMPL
CCV61	CCV61	22446	5.1612	-	ISM02.4	PPB	3/26/2019 19:37	SMPL
CCB61	CCB61	-20	-0.0611	-	ISM02.4	PPB	3/26/2019 19:40	SMPL