

**TOTAL SUSPENDED SOLIDS - SM2540D**SUPERVISOR: apatelANALYST: AparnaDate: 03/26/2019Run Number: LB101642BalanceID: WC SC-3OvenID: wet-oven#2FilterID: 15028095ThermometerID: WET-OVEN 2

TEMP1 IN:	103	03/26/2019 14:00	TEMP OUT:	104	03/26/2019 15:20
TEMP2 IN:	104	03/26/2019 16:00	TEMP OUT:	104	03/27/2019 07:10
TEMP3 IN:	105	03/27/2019 09:00	TEMP OUT:	105	03/27/2019 10:45
TEMP4 IN:	105	03/27/2019 11:30	TEMP OUT:	105	03/27/2019 13:10

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB101642BL	LB101642BL	1.4020	1.4020	100	1.4020	1.4020	1.4020	0.0000	0
2	LB101642BS	LB101642BS	1.3956	1.3956	100	1.4460	1.4460	1.4460	0.0504	504
3	K2093-17	30092-30093-5	1.4041	1.4041	10	1.4278	1.4278	1.4278	0.0237	2370
4	K2093-20	30099-30100-OW-6	1.3822	1.3822	200	1.3893	1.3893	1.3893	0.0071	35.5
5	K2099-01	SHORE-RD-BOX-CULVERT-TP-SET	1.3918	1.3918	250	1.4036	1.4036	1.4036	0.0118	47.2
6	K2099-01DUP	SHORE-RD-BOX-CULVERT-TP-SET	1.3862	1.3862	250	1.3984	1.3984	1.3984	0.0122	48.8
7	K2099-02	SHORE-RD-BOX-CULVERT-TP-SET	1.4066	1.4066	200	1.4185	1.4185	1.4185	0.0119	59.5
8	K2099-03	MH-C05-SETTLED-1H-A	1.3908	1.3908	200	1.3948	1.3948	1.3948	0.0040	20
9	K2099-04	MH-C05-SETTLED-1H-B	1.4083	1.4083	100	1.4096	1.4096	1.4096	0.0013	13
10	K2099-05	MH-C02-SETTLED-1H-A	1.4162	1.4162	100	1.4176	1.4176	1.4176	0.0014	14
11	K2099-06	MH-C02-SETTLED-1H-B	1.4038	1.4038	100	1.4055	1.4055	1.4055	0.0017	17
12	K2099-07	SHORE-RD-BOX-CULVERT-TP-SET	1.4244	1.4244	500	1.4368	1.4368	1.4368	0.0124	24.8
13	K2099-08	SHORE-RD-BOX-CULVERT-TP-SET	1.3971	1.3971	200	1.4063	1.4063	1.4063	0.0092	46
14	K2099-09	MH-C05-SETTLED-2H-A	1.3808	1.3808	100	1.3828	1.3828	1.3828	0.0020	20
15	K2099-10	MH-C05-SETTLED-2H-B	1.3938	1.3938	100	1.3960	1.3960	1.3960	0.0022	22



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 03/26/2019Run Number: LB101642BalanceID: WC SC-3OvenID: wet-oven#2FilterID: 15028095ThermometerID: WET-OVEN 2

TEMP1 IN:	<u>103</u>	<u>03/26/2019 14:00</u>	TEMP OUT:	<u>104</u>	<u>03/26/2019 15:20</u>
TEMP2 IN:	<u>104</u>	<u>03/26/2019 16:00</u>	TEMP OUT:	<u>104</u>	<u>03/27/2019 07:10</u>
TEMP3 IN:	<u>105</u>	<u>03/27/2019 09:00</u>	TEMP OUT:	<u>105</u>	<u>03/27/2019 10:45</u>
TEMP4 IN:	<u>105</u>	<u>03/27/2019 11:30</u>	TEMP OUT:	<u>105</u>	<u>03/27/2019 13:10</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	K2103-01	MH-C02-SETTLED-2H-A	1.3845	1.3845	100	1.3858	1.3858	1.3858	0.0013	13
17	K2103-02	MH-C02-SETTLED-2H-B	1.4019	1.4019	100	1.4032	1.4032	1.4032	0.0013	13
18	K2103-03	SHORE-RD-BOX-CULVERT-SETTLE	1.3928	1.3928	400	1.4015	1.4015	1.4015	0.0087	21.8
19	K2103-04	SHORE-RD-BOX-CULVERT-SETTL	1.4125	1.4125	200	1.4223	1.4223	1.4223	0.0098	49
20	K2103-05	MH-C05-SETTLED-3H-A	1.3941	1.3941	100	1.3971	1.3971	1.3971	0.0030	30
21	K2103-06	MH-C05-SETTLED-3H-B	1.3945	1.3945	100	1.3966	1.3966	1.3966	0.0021	21
22	K2103-07	MH-C02-SETTLED-3H-A	1.4146	1.4146	100	1.4158	1.4158	1.4158	0.0012	12
23	K2103-08	MH-C02-SETTLED-3H-B	1.4156	1.4156	100	1.4181	1.4181	1.4181	0.0025	25
24	K2103-08DUP	MH-C02-SETTLED-3H-BDUP	1.3980	1.3980	100	1.4004	1.4004	1.4004	0.0024	24

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)


WorkList Name : TSS032619-1

WorkList ID : 123963

Date : 03-26-2019 15:20:08

20101642

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/02/2019	Water	K2093-17	TSS	Cool 4 deg C	PSEG01	K31	30092-30093-5	03/26/2019	SM2540 D
04/02/2019	Water	K2093-20	TSS	Cool 4 deg C	PSEG01	K31	30099-30100-OW-6	03/26/2019	SM2540 D
03/28/2019	Water	K2099-01	TSS	Cool 4 deg C	TULL02	F62	SHORE-RD-BOX-CULVERT-1	03/25/2019	SM2540 D
03/28/2019	Water	K2099-02	TSS	Cool 4 deg C	TULL02	F62	SHORE-RD-BOX-CULVERT-1	03/25/2019	SM2540 D
03/28/2019	Water	K2099-03	TSS	Cool 4 deg C	TULL02	F62	MH-C05-SETTLED-1H-A	03/25/2019	SM2540 D
03/28/2019	Water	K2099-04	TSS	Cool 4 deg C	TULL02	F62	MH-C05-SETTLED-1H-B	03/25/2019	SM2540 D
03/28/2019	Water	K2099-05	TSS	Cool 4 deg C	TULL02	F62	MH-C02-SETTLED-1H-A	03/25/2019	SM2540 D
03/28/2019	Water	K2099-06	TSS	Cool 4 deg C	TULL02	F62	MH-C02-SETTLED-1H-B	03/25/2019	SM2540 D
03/28/2019	Water	K2099-07	TSS	Cool 4 deg C	TULL02	F62	SHORE-RD-BOX-CULVERT-1	03/25/2019	SM2540 D
03/28/2019	Water	K2099-08	TSS	Cool 4 deg C	TULL02	F62	SHORE-RD-BOX-CULVERT-1	03/25/2019	SM2540 D
03/28/2019	Water	K2099-09	TSS	Cool 4 deg C	TULL02	F62	MH-C05-SETTLED-2H-A	03/25/2019	SM2540 D
03/28/2019	Water	K2099-10	TSS	Cool 4 deg C	TULL02	F62	MH-C05-SETTLED-2H-B	03/25/2019	SM2540 D
03/28/2019	Water	K2103-01	TSS	Cool 4 deg C	TULL02	F62	MH-C02-SETTLED-2H-A	03/25/2019	SM2540 D
03/28/2019	Water	K2103-02	TSS	Cool 4 deg C	TULL02	F62	MH-C02-SETTLED-2H-B	03/25/2019	SM2540 D
03/28/2019	Water	K2103-03	TSS	Cool 4 deg C	TULL02	F62	SHORE-RD-BOX-CULVERT-1	03/25/2019	SM2540 D
03/28/2019	Water	K2103-04	TSS	Cool 4 deg C	TULL02	F62	SHORE-RD-BOX-CULVERT-1	03/25/2019	SM2540 D
03/28/2019	Water	K2103-05	TSS	Cool 4 deg C	TULL02	F62	MH-C05-SETTLED-3H-A	03/25/2019	SM2540 D
03/28/2019	Water	K2103-06	TSS	Cool 4 deg C	TULL02	F62	MH-C05-SETTLED-3H-B	03/25/2019	SM2540 D
03/28/2019	Water	K2103-07	TSS	Cool 4 deg C	TULL02	F62	MH-C02-SETTLED-3H-A	03/25/2019	SM2540 D
03/28/2019	Water	K2103-08	TSS	Cool 4 deg C	TULL02	F62	MH-C02-SETTLED-3H-B	03/25/2019	SM2540 D

Date/Time 03/27/19 08:00
 Received by: 
 Relinquished by: 

Date/Time 03/27/19 09:30
 Received by: 
 Relinquished by: 