



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB101644

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP73019
Turbidity std, 40NTU	WP73012
turbidity std 5 NTU	WP73016
turbidity std 20 NTU	WP73013
Turbidity std 1 NTU	WP73017
turbidity std 0 NTU	WP73006
turbidity std 10 NTU	WP73014
TURBIDITY LOD STD, 0.5NTU	WP73021

Intercept: 0.0712

Slope: 0.9668

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.270	03/27/2019	10:12
2	CAL2	1	1	1.020	03/27/2019	10:15
3	CAL3	5	1	4.680	03/27/2019	10:18
4	CAL4	10	1	9.830	03/27/2019	10:22
5	CAL5	20	1	19.300	03/27/2019	10:25
6	CAL6	40	1	38.800	03/27/2019	10:28



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB101644

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.810	10.073	03/27/2019	10:32
2	ICB		1	0.280	0.216	03/27/2019	10:36
3	CCV1	10	1	9.650	9.908	03/27/2019	10:40
4	CCB1		1	0.246	0.181	03/27/2019	10:43
5	LB101644BL		1	0.259	0.194	03/27/2019	10:46
6	K1560-09		1	0.850	0.806	03/27/2019	10:49
7	CCV2	10	1	9.910	10.177	03/27/2019	10:53
8	CCB2		1	0.275	0.211	03/27/2019	10:56