

Analysis Method: SM2510 BANALYST: AMANDEEPParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB101649

Reagent/Standard	Lot/Log #
conductivity standard	WP73030

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB101649BL		1	21.6	0.75	0.80	03/27/2019	12:15
2	LB101649BS	1413	1	22.8	1345.00	1404.00	03/27/2019	12:29
3	K1833-07		1	21.9	604.00	642.01	03/27/2019	12:35

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--