

Analysis Method: 120.1ANALYST: AMANDEEPParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB101650Meter ID: WC Conducataance Meter-1Thermometer ID: WC pH-meter Temp probe

Reagent/Standard	Lot/Log #
conductivity standard	WP73030

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB101650BL		1	21.6	0.75	0.80	03/27/2019	12:15
2	LB101650BS	1413	1	22.8	1345.00	1404.00	03/27/2019	12:29
3	K1833-08		1	21.9	606.00	644.14	03/27/2019	12:37

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--