

Analysis Method: 9050A

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB101651

Reagent/Standard	Lot/Log #
conductivity standard	WP73030

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB101651BL		1	21.6	0.75	0.80	03/27/2019	12:15
2	LB101651BS	1413	1	22.8	1345.00	1404.00	03/27/2019	12:29
3	K1833-09		1	21.9	605.00	643.08	03/27/2019	12:41

Note : Conductivity result is corrected to 25°C

$$\text{Result} = \frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$$