



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101656

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/27/2019	12:40
2	CAL2	1	Water	NA	NA	20.1	7.01	03/27/2019	12:41
3	CAL3	1	Water	NA	NA	20.1	10.02	03/27/2019	12:43
4	ICV	1	Water	NA	NA	20.1	6.99	03/27/2019	12:45
5	CCV1	1	Water	NA	NA	20.1	2.01	03/27/2019	12:46
6	K2093-18	1	Water	NA	NA	20.1	8.43	03/27/2019	13:10
7	K2093-18DUP	1	Water	NA	NA	20.2	8.44	03/27/2019	13:11
8	K2093-21	1	Water	NA	NA	20.6	5.34	03/27/2019	13:12
9	CCV2	1	Water	NA	NA	20.1	12.02	03/27/2019	13:38

LB 101656

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K2093

WorkList ID : 124014

Date : 03-27-2019 12:31:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2019	Water	K2093-18-A	Corrosivity	Cool 4 deg C	PSEG01	K31	30092-30093-5	03/26/2019	9040C
03/26/2019	Water	K2093-21-A	Corrosivity	Cool 4 deg C	PSEG01	K31	30099-30100-OW-6	03/26/2019	9040C

Date/Time

03-27-19 12:35

Received by:

JO

Relinquished by:

JO

Date/Time

03-27-19 14:00

Received by:

JO

Relinquished by:

JO