

LB101672


INSTRUMENT ID: CV1

ISM02.4

Linear

 μ Abs.:

52057



A= 0.0000e+000

B= 1.9259e-004

C= -5.3232e-002

Rho= 0.9999574

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	-0.011	-0.011	221	0.000	221					
0.2	0.200	0.209	0.009	1364	0.0 %	1364					5
2.5	2.500	2.458	-0.042	13037	0.0 %	13037					-2
5.0	5.000	5.064	0.064	26573	0.0 %	26573					1
7.5	7.500	7.507	0.007	39254	0.0 %	39254					0
10.0	10.000	9.972	-0.028	52057	0.0 %	52057					0

RS 03-27-19

LB101672
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	221	-		0 ISM02.4	PPB	3/27/2019 15:29	Std
	0.2 S0.2	1364	-		0.2 ISM02.4	PPB	3/27/2019 15:31	Std
	2.5 S2.5	13037	-		2.5 ISM02.4	PPB	3/27/2019 15:34	Std
	5 S5.0	26573	-		5 ISM02.4	PPB	3/27/2019 15:36	Std
	7.5 S7.5	39254	-		7.5 ISM02.4	PPB	3/27/2019 15:38	Std
	10 S10.0	52057	-		10 ISM02.4	PPB	3/27/2019 15:40	Std
ICV	ICV	21120	4.0143 -		ISM02.4	PPB	3/27/2019 15:45	SMPL
ICB	ICB	-10	-0.0552 -		ISM02.4	PPB	3/27/2019 15:48	SMPL
CCV065	CCV065	26733	5.0953 -		ISM02.4	PPB	3/27/2019 15:50	SMPL
CCB065	CCB065	122	-0.0297 -		ISM02.4	PPB	3/27/2019 15:52	SMPL
PB118260BL	PBS260	42	-0.0451 -		ISM02.4	PPB	3/27/2019 15:54	SMPL
K2046-01	MBF5T1	1061	0.1511 -		ISM02.4	PPB	3/27/2019 15:57	SMPL
K2046-02	MBF5T2	2606	0.4487 -		ISM02.4	PPB	3/27/2019 15:59	SMPL
K2046-03	MBF5T3	13509	2.5485 -		ISM02.4	PPB	3/27/2019 16:01	SMPL
K2046-04	MBF5T4	3317	0.5856 -		ISM02.4	PPB	3/27/2019 16:03	SMPL
K2046-05	MBF5T5	2952	0.5153 -		ISM02.4	PPB	3/27/2019 16:06	SMPL
K2046-06	MBF5T6	6990	1.293 -		ISM02.4	PPB	3/27/2019 16:08	SMPL
K2046-07	MBF5T7	4051	0.727 -		ISM02.4	PPB	3/27/2019 16:10	SMPL
K2046-08	MBF5T8	2989	0.5224 -		ISM02.4	PPB	3/27/2019 16:13	SMPL
K2046-09	MBF5W2	698	0.0812 -		ISM02.4	PPB	3/27/2019 16:15	SMPL
K2046-10	MBF5W3	1277	0.1927 -		ISM02.4	PPB	3/27/2019 16:17	SMPL
K2046-11	MBF5W4	3819	0.6823 -		ISM02.4	PPB	3/27/2019 16:19	SMPL
K2046-12	MBF5W5	848	0.1101 -		ISM02.4	PPB	3/27/2019 16:22	SMPL
K2046-13	MBF5W6	2305	0.3907 -		ISM02.4	PPB	3/27/2019 16:24	SMPL
K2046-14	MBF5W6D	2329	0.3953 -		ISM02.4	PPB	3/27/2019 16:26	SMPL
K2046-15	MBF5W6S	18756	3.559 -		ISM02.4	PPB	3/27/2019 16:36	SMPL
K2046-16	MBF5W7	672	0.0762 -		ISM02.4	PPB	3/27/2019 16:38	SMPL
PB118261BL	PBS261	302	0.0049 -		ISM02.4	PPB	3/27/2019 16:41	SMPL
K2021-01	MBF5R8	15710	2.9724 -		ISM02.4	PPB	3/27/2019 16:43	SMPL
K2021-02	MBF5R8D	15482	2.9285 -		ISM02.4	PPB	3/27/2019 16:45	SMPL
CCV066	CCV066	26873	5.1223 -		ISM02.4	PPB	3/27/2019 16:48	SMPL
CCB066	CCB066	18	-0.0498 -		ISM02.4	PPB	3/27/2019 16:50	SMPL
K2021-03	MBF5R8S	33884	6.4725 -		ISM02.4	PPB	3/27/2019 16:52	SMPL
K2021-04	MBF5R9	2213	0.373 -		ISM02.4	PPB	3/27/2019 16:54	SMPL
K2021-05	MBF5S0	2115	0.3541 -		ISM02.4	PPB	3/27/2019 16:57	SMPL
K2021-06	MBF5S1	592	0.0608 -		ISM02.4	PPB	3/27/2019 16:59	SMPL
K2021-07	MBF5S2	688	0.0793 -		ISM02.4	PPB	3/27/2019 17:01	SMPL
K2021-08	MBF5S3	2135	0.3579 -		ISM02.4	PPB	3/27/2019 17:03	SMPL
K2021-09	MBF5S4	17537	3.3242 -		ISM02.4	PPB	3/27/2019 17:06	SMPL
K2021-10	MBF5S5	4502	0.8138 -		ISM02.4	PPB	3/27/2019 17:08	SMPL
K2021-11	MBF5S6	9325	1.7427 -		ISM02.4	PPB	3/27/2019 17:10	SMPL
K2021-12	MBF5S7	1701	0.2744 -		ISM02.4	PPB	3/27/2019 17:12	SMPL
K2021-13	MBF5S8	3184	0.56 -		ISM02.4	PPB	3/27/2019 17:15	SMPL
K2021-14	MBF5S9	3713	0.6619 -		ISM02.4	PPB	3/27/2019 17:17	SMPL
K2021-15	MBF5T0	15311	2.8955 -		ISM02.4	PPB	3/27/2019 17:19	SMPL
K2021-16	MBF5T9	1502	0.236 -		ISM02.4	PPB	3/27/2019 17:21	SMPL

LB101672

INSTRUMENT ID : CV1

K2021-17	MBF5W0	1700	0.2742 -	ISM02.4	PPB	3/27/2019 17:24 SMPL
K2021-18	MBF5W1	2240	0.3782 -	ISM02.4	PPB	3/27/2019 17:26 SMPL
K2021-19	MBF5X6	3116	0.5469 -	ISM02.4	PPB	3/27/2019 17:28 SMPL
K2021-20	MBF5X8	3854	0.689 -	ISM02.4	PPB	3/27/2019 17:31 SMPL
CCV067	CCV067	25333	4.8257 -	ISM02.4	PPB	3/27/2019 17:33 SMPL
CCB067	CCB067	-25	-0.058 -	ISM02.4	PPB	3/27/2019 17:35 SMPL