



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101674

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	03/27/2019	14:15
2	K2093-03		1	8	>100.0	>212.0	>212.0	03/27/2019	14:32
3	K2093-03DUP		1	8	>100.0	>212.0	>212.0	03/27/2019	14:50
4	K2093-15		1	9	>100.0	>212.0	>212.0	03/27/2019	15:20
5	K2093-18		1	9	>100.0	>212.0	>212.0	03/27/2019	15:45
6	K2093-21		1	7	>100.0	>212.0	>212.0	03/27/2019	16:10
7	K2110-02		1	8	>100.0	>212.0	>212.0	03/27/2019	16:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB101674

WorkList Name : FP32719

WorkList ID : 124000

Date : 03-27-2019 10:43:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/04/2019	Water	K2093-03	Flash Point	Cool 4 deg C	PSEG01	K31	DL-2-COMP	03/26/2019	1010A
04/04/2019	Water	K2093-15	Flash Point	Cool 4 deg C	PSEG01	K31	30063-3	03/26/2019	1010A
04/04/2019	Water	K2093-18	Flash Point	Cool 4 deg C	PSEG01	K31	30092-30093-5	03/26/2019	1010A
04/04/2019	Water	K2093-21	Flash Point	Cool 4 deg C	PSEG01	K31	30099-30100-OW-6	03/26/2019	1010A
04/04/2019	Water	K2110-02	Flash Point	Cool 4 deg C	PSEG01	K31	BUR-COMP	03/26/2019	1010A

Date/Time 03/27/19 13:50
 Received by: KS
 Relinquished by: CP

Date/Time 03/27/19 16:40
 Received by: CP
 Relinquished by: KS