



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 3/28/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:30
In Date: 03/27/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 03/28/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101677

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1697-13	SU-02-032719-A	1	1.14	9.71	8.92	90.8
K1697-15	SU-02-032719-B	2	1.16	9.87	8.74	87.0
K1982-03	EXT-028-SW004-032519	3	1.19	9.84	8.76	87.5
K1982-04	EXT-028-SW005-032519	4	1.1	9.88	8.66	86.1
K1982-05	EXT-028-SW008-032519	5	1.15	9.69	8.54	86.5
K1982-06	EXT-028-SW009-032519	6	1.18	9.98	9.11	90.1
K1982-07	EXT-028-SW010-032519	7	1.15	9.94	8.84	87.5
K1982-08	EXT-027-SW011-032519	8	1.13	9.71	8.64	87.5
K1982-09	EXT-027-SW012-032519	9	1.14	9.86	8.69	86.6
K1982-10	EXT-027-SW013-032519	10	1.16	9.54	8.39	86.3
K1982-11	EXT-027-SW014-032519	11	1.17	9.94	9.35	93.3
K1982-12	EXT-027-SW024-032519	12	1.12	9.83	8.45	84.2
K1982-13	PST-027-B110-032519	13	1.14	9.72	8.9	90.4
K1982-14	PST-027-B114-032519	14	1.11	9.76	8.73	88.1
K2111-01	S5 (5-15) COMP	15	1.18	9.88	9.44	94.9
K2111-03	S5 (17-25) COMP	16	1.1	9.84	9.23	93.0
K2111-05	S7A (4-20) COMP	17	1.18	9.97	9.38	93.3
K2111-07	S5 (15)	18	1.15	9.82	8.61	86.0
K2111-09	S5 (23)	19	1.16	9.75	8.71	87.9
K2111-11	S7A (20)	20	1.19	9.97	9.44	94.0
K2112-01	EPH-4-8-WC-001	21	1.17	9.67	7.7	76.8
K2112-02	EPH-4-8-WC-002	22	1.15	9.95	8.38	82.2
K2112-03	DR-WC-001	23	1.15	9.97	6.74	63.4
K2112-06	EPH-4-8-WC-001	24	1.17	9.67	7.7	76.8
K2112-07	EPH-4-8-WC-002	25	1.15	9.95	8.38	82.2
K2112-08	DR-WC-001	26	1.15	9.97	6.74	63.4
K2125-01	CONCRETE-COMP-A	27	1.00	2.00	2.00	100.0
K2125-02	CONCRETE-COMP-B	28	1.00	2.00	2.00	100.0
K2125-03	SOIL-COMP	29	1.16	9.97	9.00	89.0
K2129-01	PE-4D_8.0-8.5	30	1.12	9.94	8.05	78.6
K2129-02	PE-7D_8.5-9.0	31	1.17	9.94	8.73	86.2
K2130-01	FK-GF-02-034-A	32	1.14	9.8	8.54	85.5
K2130-02	FK-GF-02-034-A	33	1.12	9.97	8.78	86.6
K2130-03	FK-GF-02-034-B	34	1.12	9.98	8.88	87.6
K2130-04	FK-GF-02-034-B	35	1.12	9.51	8.29	85.5
K2131-01	FK-GF-02-035-A	36	1.1	9.97	8.84	87.3
K2131-02	FK-GF-02-035-A	37	1.13	9.96	9.01	89.2
K2131-03	FK-GF-02-035-A	38	1.13	9.98	8.88	87.6
K2131-04	FK-GF-02-035-A	39	1.13	9.9	8.66	85.9



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 3/28/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:30
In Date: 03/27/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 03/28/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101677

<u>Lab ID</u>	<u>Client Sample ID</u>	<u>Dish#</u>	<u>Dish Wt(g)</u> <u>(A)</u>	<u>Dish +</u> <u>Sample Wt(g)</u> <u>(B)</u>	<u>Dish + Dry</u> <u>Sample Wt(g)</u> <u>(C)</u>	<u>% Solid</u>
K2132-01	FK-GF-02-036-A	40	1.14	9.65	8.26	83.7
K2132-02	FK-GF-02-036-A	41	1.12	9.71	8.53	86.3
K2132-03	FK-GF-02-036-A	42	1.12	9.98	9.00	88.9
K2132-04	FK-GF-02-036-A	43	1.14	9.97	8.88	87.7

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$